

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112222\
 Data File : PS021217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2022 14:26
 Operator : AR\AJ
 Sample : N5678-01MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 250FB-BP-BACKFILL-01MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/23/2022
 Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 23:32:24 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111622.M
 Quant Title : 8080.M
 QLast Update : Thu Nov 17 00:46:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.406	6.840	519.7E6	315.4E6	239.021	208.586m
Target Compounds							
1) T	Dalapon	2.495	2.567	473.4E6	446.1E6	170.097m	221.302m#
2) T	3,5-DICHL...	5.774	6.014	554.2E6	390.3E6	175.281	167.927
3) T	4-Nitroph...	6.247	6.453	303.4E6	163.0E6	224.601	185.330
5) T	DICAMBA	6.550	7.006	1880.6E6	1345.8E6	205.028	195.519
6) T	MCPD	6.724	7.099	352.4E6	68708930	52.163m	13.174 #
7) T	MCPA	6.827	7.310	157.6E6	136.3E6	16.398	18.484
8) T	DICHLORPROP	7.150	7.632	356.7E6	303.8E6	151.657	177.917
9) T	2,4-D	7.354	7.924	586.9E6	419.6E6	206.883	206.341
10) T	Pentachlo...	7.582	8.353	3438.6E6	2359.9E6	104.942	99.442
11) T	2,4,5-TP ...	8.110	8.715	1878.7E6	1335.0E6	135.677	134.186
12) T	2,4,5-T	8.371	9.093	2232.9E6	1457.3E6	158.603	157.028
13) T	2,4-DB	8.882	9.602	234.8E6	124.6E6	102.747m	106.413
14) T	DINOSEB	9.936	9.929	195.2E6	67284865	17.015	9.589 #
15) T	Picloram	9.776	10.906	4500.9E6	2950.8E6	223.819	228.590
16) T	DCPA	10.202	10.858	5310.5E6	3190.0E6	280.947	266.428

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112222\
Data File : PS021217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2022 14:26
Operator : AR\AJ
Sample : N5678-01MS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
250FB-BP-BACKFILL-01MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/23/2022
Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 23:32:24 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111622.M
Quant Title : 8080.M
QLast Update : Thu Nov 17 00:46:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

