

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112420\
 Data File : PS013301.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Nov 2020 11:25
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 24 11:48:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 03 03:59:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.208	6.912	699.0E6	305.4E6	781.926	784.921
Target Compounds							
1) T	Dalapon	3.032	2.624	710.2E6	354.6E6	671.269	669.032
2) T	3,5-DICHL...	6.425	6.066	892.1E6	445.8E6	724.365	724.635
3) T	4-Nitroph...	7.043	6.525	308.5E6	174.4E6	723.270	732.145
5) T	DICAMBA	7.398	7.078	2635.5E6	1222.4E6	737.984	736.397
6) T	MCPD	7.565	7.173	198.7E6	95603064	74.640	72.230
7) T	MCPA	7.714	7.388	276.9E6	148.4E6	72.511	69.753
8) T	DICHLORPROP	8.089	7.704	652.7E6	325.8E6	728.904	729.033
9) T	2,4-D	8.321	8.001	715.4E6	390.0E6	757.277	734.829
10) T	Pentachlo...	8.627	8.425	9551.4E6	4098.3E6	742.135	741.608
11) T	2,4,5-TP ...	9.176	8.787	3818.3E6	1735.3E6	743.506	748.780
12) T	2,4,5-T	9.463	9.167	3909.1E6	1686.7E6	738.673	761.477
13) T	2,4-DB	10.017	9.674	624.8E6	249.5E6	741.669	768.708
14) T	DINOSEB	11.169	10.002	2963.8E6	1254.6E6	718.750	744.180
15) T	Picloram	11.002	10.994	4871.6E6	2563.5E6	740.851	710.209
16) T	DCPA	11.435	10.928	4832.3E6	2265.8E6	744.099	754.713

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112420\
Data File : PS013301.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2020 11:25
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 11:48:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
Quant Title : 8080.M
QLast Update : Tue Nov 03 03:59:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

