

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112421\
 Data File : PS017522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2021 00:40
 Operator : AJ\MA
 Sample : PB141006BSD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB141006BSD

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 11/26/2021
 Supervised By :mohammad ahmed 11/26/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 01:13:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111821.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 19 21:46:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.146	6.735	809.8E6	322.2E6	515.097	543.721
Target Compounds							
1) T	Dalapon	2.989	2.500	825.7E6	232.9E6	352.736	343.880m
2) T	3,5-DICHL...	6.370	5.934	1025.4E6	403.8E6	467.597	452.932
3) T	4-Nitroph...	6.976	6.362	345.7E6	124.3E6	413.335	380.696
5) T	DICAMBA	7.334	6.896	2667.8E6	1150.1E6	437.088	442.333
6) T	MCPD	7.502	6.990	202.2E6	92725025	43.572	44.091
7) T	MCPA	7.649	7.200	290.7E6	141.1E6	44.357	48.159
8) T	DICHLORPROP	8.025	7.512	670.4E6	312.0E6	436.574	473.810
9) T	2,4-D	8.254	7.803	865.1E6	383.9E6	462.977	492.124
10) T	Pentachlo...	8.558	8.226	10613.5E6	4192.0E6	502.531	499.352
11) T	2,4,5-TP ...	9.107	8.586	4094.4E6	1680.6E6	491.335	474.000m
12) T	2,4,5-T	9.392	8.964	3831.6E6	1514.0E6	453.551	455.272
13) T	2,4-DB	9.943	9.468	599.4E6	194.9E6	352.262	457.153 #
14) T	DINOSEB	11.092	9.793	2759.5E6	995.8E6	443.125m	422.369
15) T	Picloram	10.918	10.774	4394.0E6	1545.5E6	360.389	354.171
16) T	DCPA	11.360	10.723	4324.3E6	2010.5E6	404.932	456.271

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112421\
Data File : PS017522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2021 00:40
Operator : AJ\MA
Sample : PB141006BSD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB141006BSD

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 11/26/2021
Supervised By :mohammad ahmed 11/26/2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 01:13:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111821.M
Quant Title : 8080.M
QLast Update : Fri Nov 19 21:46:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

