

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112421\
 Data File : PS017525.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2021 01:53
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 02:44:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111821.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 19 21:46:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.146	6.736	1277.5E6	485.6E6	812.572	819.552
Target Compounds							
1) T	Dalapon	2.987	2.495	1471.2E6	430.6E6	628.456	635.862
2) T	3,5-DICHL...	6.371	5.935	1661.2E6	684.8E6	757.538	768.186
3) T	4-Nitroph...	6.976	6.362	634.1E6	246.9E6	758.124	756.362
5) T	DICAMBA	7.334	6.896	4713.8E6	2012.2E6	772.309	773.909
6) T	MCPD	7.505	6.993	346.4E6	163.1E6	74.666	77.567
7) T	MCPA	7.653	7.205	505.5E6	223.6E6	77.122	76.339
8) T	DICHLORPROP	8.025	7.512	1172.8E6	514.0E6	763.722	780.631
9) T	2,4-D	8.254	7.803	1415.9E6	600.6E6	757.761	770.063
10) T	Pentachlo...	8.558	8.226	16349.9E6	6407.7E6	774.135	763.297
11) T	2,4,5-TP ...	9.108	8.587	6485.3E6	2760.5E6	778.250	778.566
12) T	2,4,5-T	9.392	8.963	6614.3E6	2597.3E6	782.947	781.040
13) T	2,4-DB	9.941	9.468	1307.9E6	333.5E6	768.720	782.270
14) T	DINOSEB	11.091	9.793	4780.8E6	1862.3E6	767.704	789.845
15) T	Picloram	10.915	10.773	8013.0E6	2840.6E6	657.215	650.986
16) T	DCPA	11.360	10.722	8344.3E6	3434.3E6	781.371	779.401

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112421\
Data File : PS017525.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2021 01:53
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 02:44:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111821.M
Quant Title : 8080.M
QLast Update : Fri Nov 19 21:46:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

