

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112520\
 Data File : PS013330.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2020 19:01
 Operator : DD\AJ
 Sample : L4826-01MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB-2020-WS-000-04MS

Manual Integrations
 APPROVED

Ankita
 12/9/2020 3:22:36 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 07:47:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 03 03:59:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.208	6.910	458.2E6	237.6E6	512.518	610.753
Target Compounds							
1) T	Dalapon	3.030	2.624	629.0E6	394.3E6	594.553	744.028 #
2) T	3,5-DICHL...	6.423	6.064	754.9E6	351.0E6	612.918	570.621
5) T	DICAMBA	7.397	7.077	2073.9E6	978.3E6	580.724	589.327
6) T	MCP	7.564	7.171	163.2E6	78298136	61.306	59.156
7) T	MCPA	7.712	7.385	298.1E6	122.1E6	78.067m	57.407m#
8) T	DICHLORPROP	8.088	7.702	536.3E6	273.4E6	598.895	611.776
9) T	2,4-D	8.319	7.997	647.5E6	366.7E6	685.362	691.018
10) T	Pentachlo...	8.626	8.423	7737.3E6	3349.8E6	601.182	606.164
11) T	2,4,5-TP ...	9.174	8.783	3294.4E6	1741.8E6	641.484	751.585
12) T	2,4,5-T	9.460	9.164	3277.0E6	1461.1E6	619.224	659.650
13) T	2,4-DB	10.013	9.670	367.0E6	193.3E6	435.582m	595.452m#
14) T	DINOSEB	11.168	10.000	2113.2E6	898.2E6	512.466	532.802
15) T	Picloram	10.999	10.991	3493.6E6	1830.7E6	531.287	507.207
16) T	DCPA	11.434	10.926	3861.2E6	1862.5E6	594.571	620.365

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112520\
Data File : PS013330.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2020 19:01
Operator : DD\AJ
Sample : L4826-01MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB-2020-WS-000-04MS

Manual Integrations
APPROVED

Ankita
12/9/2020 3:22:36 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 07:47:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
Quant Title : 8080.M
QLast Update : Tue Nov 03 03:59:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

