

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112524\
 Data File : PS028618.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2024 11:51
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/26/2024
 Supervised By :Ankita Jodhani 11/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 05:24:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 13 13:41:03 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.226	7.727	1943.6E6	940.2E6	760.440m	727.558m
Target Compounds							
1) T	Dalapon	2.635	2.695	2407.6E6	1621.5E6	716.668	670.081
2) T	3,5-DICHL...	6.400	6.684	2788.1E6	1333.6E6	747.223	690.117
3) T	4-Nitroph...	7.025	7.254	1242.3E6	599.4E6	731.460m	694.194
5) T	DICAMBA	7.414	7.926	8451.7E6	4176.1E6	772.411	732.134
6) T	MCPD	7.597	8.030	535.1E6	272.3E6	76.061	63.315
7) T	MCPA	7.747	8.274	742.6E6	385.7E6	72.772	62.276
8) T	DICHLORPROP	8.121	8.642	2220.2E6	1012.7E6	752.251	693.329
9) T	2,4-D	8.351	8.972	2392.8E6	1102.8E6	748.483	704.435
10) T	Pentachlo...	8.650	9.500	33968.1E6	15484.4E6	813.879	772.469
11) T	2,4,5-TP ...	9.228	9.877	13426.9E6	6323.8E6	772.961	749.413
12) T	2,4,5-T	9.520	10.296	13722.0E6	6126.3E6	772.005	730.861
13) T	2,4-DB	10.095	10.863	2015.3E6	696.6E6	739.401	664.869
14) T	DINOSEB	11.305	11.242	10646.0E6	3857.7E6	746.680	697.785
15) T	Picloram	11.113	12.335	19948.2E6	7862.0E6	708.395	694.583
16) T	DCPA	11.598	12.285	20346.7E6	7504.2E6	773.245	773.754

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112524\
Data File : PS028618.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2024 11:51
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/26/2024
Supervised By :Ankita Jodhani 11/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 05:24:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111324.M
Quant Title : 8080.M
QLast Update : Wed Nov 13 13:41:03 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

