

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS112621\  
 Data File : PS017529.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 26 Nov 2021 10:33  
 Operator : AJ\MA  
 Sample : HSTDCCC750  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :  
 HSTDCCC750

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Nov 26 21:39:17 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS111821.M  
 Quant Title : 8080.M  
 QLast Update : Fri Nov 19 21:46:29 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1    | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|--------------|--------|--------|-----------|----------|---------|---------|
| -----                       |              |        |        |           |          |         |         |
| System Monitoring Compounds |              |        |        |           |          |         |         |
| 4) S                        | 2,4-DCAA     | 7.145  | 6.736  | 1215.4E6  | 457.4E6  | 773.057 | 771.823 |
| Target Compounds            |              |        |        |           |          |         |         |
| 1) T                        | Dalapon      | 2.987  | 2.493  | 1442.3E6  | 415.8E6  | 616.136 | 614.010 |
| 2) T                        | 3,5-DICHL... | 6.370  | 5.935  | 1567.6E6  | 643.7E6  | 714.841 | 722.077 |
| 3) T                        | 4-Nitroph... | 6.975  | 6.362  | 592.2E6   | 230.0E6  | 707.976 | 704.605 |
| 5) T                        | DICAMBA      | 7.334  | 6.896  | 4494.4E6  | 1908.0E6 | 736.360 | 733.846 |
| 6) T                        | MCP          | 7.504  | 6.992  | 324.0E6   | 154.1E6  | 69.835  | 73.267  |
| 7) T                        | MCPA         | 7.653  | 7.203  | 477.0E6   | 210.2E6  | 72.779  | 71.761  |
| 8) T                        | DICHLORPROP  | 8.025  | 7.512  | 1122.5E6  | 480.1E6  | 730.986 | 729.239 |
| 9) T                        | 2,4-D        | 8.254  | 7.803  | 1340.7E6  | 565.0E6  | 717.501 | 724.306 |
| 10) T                       | Pentachlo... | 8.559  | 8.226  | 15696.1E6 | 6160.9E6 | 743.180 | 733.891 |
| 11) T                       | 2,4,5-TP ... | 9.108  | 8.587  | 6185.8E6  | 2616.0E6 | 742.305 | 737.821 |
| 12) T                       | 2,4,5-T      | 9.392  | 8.963  | 6225.0E6  | 2423.0E6 | 736.870 | 728.619 |
| 13) T                       | 2,4-DB       | 9.942  | 9.468  | 1143.6E6  | 302.7E6  | 672.161 | 710.030 |
| 14) T                       | DINOSEB      | 11.092 | 9.793  | 4499.9E6  | 1700.9E6 | 722.591 | 721.392 |
| 15) T                       | Picloram     | 10.917 | 10.773 | 7450.2E6  | 2574.1E6 | 611.054 | 589.897 |
| 16) T                       | DCPA         | 11.360 | 10.722 | 7893.1E6  | 3237.6E6 | 739.124 | 734.747 |
| -----                       |              |        |        |           |          |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS112621\  
Data File : PS017529.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 26 Nov 2021 10:33  
Operator : AJ\MA  
Sample : HSTDCCC750  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
ECD\_S  
ClientSampleId :  
HSTDCCC750

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 26 21:39:17 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS111821.M  
Quant Title : 8080.M  
QLast Update : Fri Nov 19 21:46:29 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

