

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112621\
 Data File : PS017535.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2021 13:07
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 21:42:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111821.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 19 21:46:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.144	6.735	1212.2E6	456.1E6	771.024	769.743
Target Compounds							
1) T	Dalapon	2.989	2.493	1437.2E6	416.4E6	613.969	614.876
2) T	3,5-DICHL...	6.370	5.935	1567.3E6	643.2E6	714.706	721.527
3) T	4-Nitroph...	6.975	6.362	582.4E6	229.0E6	696.321	701.716
5) T	DICAMBA	7.333	6.895	4478.3E6	1908.2E6	733.722	733.909
6) T	MCPD	7.503	6.992	318.0E6	154.0E6	68.532	73.221
7) T	MCPA	7.651	7.203	475.2E6	210.2E6	72.502	71.753
8) T	DICHLORPROP	8.024	7.512	1120.5E6	480.9E6	729.688	730.326
9) T	2,4-D	8.254	7.803	1339.9E6	563.7E6	717.084	722.661
10) T	Pentachlo...	8.557	8.226	15682.2E6	6128.6E6	742.524	730.041
11) T	2,4,5-TP ...	9.107	8.586	6158.6E6	2607.4E6	739.043	735.392
12) T	2,4,5-T	9.392	8.963	6197.9E6	2413.2E6	733.659	725.679
13) T	2,4-DB	9.941	9.468	1127.3E6	303.5E6	662.580	711.859
14) T	DINOSEB	11.091	9.792	4487.3E6	1695.5E6	720.571	719.137
15) T	Picloram	10.917	10.773	7002.2E6	2408.7E6	574.309	552.002
16) T	DCPA	11.359	10.722	7876.3E6	3227.6E6	737.547	732.482

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112621\
Data File : PS017535.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2021 13:07
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 21:42:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS111821.M
Quant Title : 8080.M
QLast Update : Fri Nov 19 21:46:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

