

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112624\
 Data File : PS028632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2024 12:48
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDICC200

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 11/27/2024

Supervised By :Ankita Jodhani 11/27/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 14:01:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 26 13:56:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	7.725	596.2E6	282.8E6	209.342m	206.925
Target Compounds							
1) T	Dalapon	2.630	2.687	531.5E6	407.6E6	183.455	185.641
2) T	3,5-DICHL...	6.399	6.681	800.4E6	379.1E6	200.243	190.612
3) T	4-Nitroph...	7.025	7.252	336.7E6	176.5E6	191.472	190.143
5) T	DICAMBA	7.412	7.923	2313.7E6	1110.1E6	195.684	182.160
6) T	MCPD	7.591	8.022	118.7E6	76224788	16.962	17.997
7) T	MCPA	7.740	8.265	186.0E6	120.6E6	18.166	19.223
8) T	DICHLORPROP	8.120	8.639	669.8E6	311.0E6	205.162	192.771
9) T	2,4-D	8.351	8.969	722.5E6	335.1E6	203.914	194.718
10) T	Pentachlo...	8.648	9.495	9682.2E6	4480.0E6	201.091	191.001
11) T	2,4,5-TP ...	9.227	9.873	3851.6E6	1833.5E6	201.635	190.020
12) T	2,4,5-T	9.519	10.293	3956.2E6	1799.0E6	201.698	190.769
13) T	2,4-DB	10.094	10.859	728.6E6	220.6E6	199.916	192.399
14) T	DINOSEB	11.303	11.238	3064.6E6	1182.6E6	197.062	187.415
15) T	Picloram	11.112	12.331	5805.4E6	2191.0E6	192.141	174.833
16) T	DCPA	11.596	12.281	5832.7E6	2127.2E6	202.702	190.123

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112624\
Data File : PS028632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2024 12:48
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/27/2024
Supervised By :Ankita Jodhani 11/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 14:01:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
Quant Title : 8080.M
QLast Update : Tue Nov 26 13:56:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

