

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112624\
 Data File : PS028651.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2024 22:01
 Operator : AR\AJ
 Sample : P4961-01MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 C0RB8MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/27/2024
 Supervised By :Ankita Jodhani 11/27/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 00:23:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 26 14:44:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.226	7.723	491.1E6	239.7E6	183.087m	177.934
Target Compounds							
1) T	Dalapon	2.632	2.692	682.4E6	870.5E6	235.173	398.636 #
2) T	3,5-DICHL...	6.400	6.680	1262.5E6	605.7E6	329.661	308.366
5) T	DICAMBA	7.414	7.922	4332.6E6	2375.9E6	377.383	385.939
6) T	MCPD	7.593	8.022	238.9E6	153.2E6	32.731m	35.561m
7) T	MCPA	7.742	8.267	342.9E6	263.3E6	33.162m	42.420 #
8) T	DICHLORPROP	8.121	8.638	1171.1E6	599.3E6	376.177	374.893
9) T	2,4-D	8.351	8.967	1553.3E6	802.4E6	460.627	474.498
10) T	Pentachlo...	8.650	9.494	9820.4E6	4847.7E6	222.723	211.772
11) T	2,4,5-TP ...	9.227	9.875	7937.3E6	7466.5E6	434.213m	782.448m#
12) T	2,4,5-T	9.519	10.291	7900.1E6	3926.0E6	421.213m	421.687
13) T	2,4-DB	10.095	10.857	1007.0E6	416.7E6	284.334	363.367 #
14) T	DINOSEB	11.303	11.235	1634.1E6	624.6E6	108.753	99.514
15) T	Picloram	11.110	12.328	10964.8E6	4659.8E6	368.729	361.486
16) T	DCPA	11.597	12.278	11437.1E6	4800.9E6	415.969	432.626

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112624\
Data File : PS028651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2024 22:01
Operator : AR\AJ
Sample : P4961-01MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
CORB8MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/27/2024
Supervised By :Ankita Jodhani 11/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 00:23:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
Quant Title : 8080.M
QLast Update : Tue Nov 26 14:44:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

