

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112719\
 Data File : PS007874.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Nov 2019 19:10
 Operator : SM\AJ
 Sample : K5675-11MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TR-05-112619MSD

Manual Integrations
 APPROVED

mohammad
 12/2/2019 2:01:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 28 02:48:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 06 02:45:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.445	5.168	228.9E6	124.3E6	166.187	176.916
Target Compounds							
1) T	Dalapon	2.021	1.734	961.4E6	546.6E6	531.419	460.365m
2) T	3,5-DICHL...	4.815	4.476	778.2E6	371.3E6	436.162m	411.011m
5) T	DICAMBA	5.590	5.291	2673.2E6	1373.9E6	525.452	519.162
6) T	MCP	5.739	5.373	195.7E6	99613252	51.347	51.276
7) T	MCPA	5.870	5.547	320.2E6	156.0E6	58.165	51.912m
8) T	DICHLORPROP	6.211	5.817	737.0E6	388.2E6	583.083	591.923
9) T	2,4-D	6.424	6.086	907.8E6	512.7E6	647.260	635.818
10) T	Pentachlo...	6.682	6.448	5220.1E6	2627.4E6	275.453	268.308
11) T	2,4,5-TP ...	7.214	6.809	4707.2E6	5097.7E6	626.372	1301.812 #
12) T	2,4,5-T	7.484	7.170	4743.2E6	2225.1E6	579.425	602.524
13) T	2,4-DB	8.013	7.654	570.9E6	292.5E6	533.876	598.314
14) T	DINOSEB	9.095	7.962	1125.5E6	491.2E6	187.411	168.980
15) T	Picloram	8.931	8.895	5879.8E6	3004.8E6	529.566	512.852
16) T	DCPA	9.367	8.852	3983.7E6	2028.1E6	406.163	388.137

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112719\
Data File : PS007874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Nov 2019 19:10
Operator : SM\AJ
Sample : K5675-11MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TR-05-112619MSD

Manual Integrations
APPROVED

mohammad
12/2/2019 2:01:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 28 02:48:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
Quant Title : 8080.M
QLast Update : Wed Nov 06 02:45:11 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

