

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112720\
 Data File : PS013344.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Nov 2020 10:01
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 11:12:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 03 03:59:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.208	6.911	697.4E6	309.1E6	780.149	794.405
Target Compounds							
1) T	Dalapon	3.033	2.625	685.3E6	348.1E6	647.769	656.791
2) T	3,5-DICHL...	6.425	6.066	881.5E6	447.9E6	715.776	728.126
3) T	4-Nitroph...	7.043	6.524	312.9E6	178.1E6	733.709	747.897
5) T	DICAMBA	7.398	7.078	2627.1E6	1243.3E6	735.640	748.972
6) T	MCPD	7.566	7.172	198.6E6	98203050	74.597	74.194
7) T	MCPA	7.714	7.387	278.1E6	150.0E6	72.839	70.485
8) T	DICHLORPROP	8.089	7.703	650.6E6	327.5E6	726.561	732.720
9) T	2,4-D	8.321	8.000	727.0E6	398.5E6	769.488	750.965
10) T	Pentachlo...	8.627	8.424	9474.7E6	4137.1E6	736.169	748.619
11) T	2,4,5-TP ...	9.176	8.785	3782.4E6	1746.8E6	736.518	753.735
12) T	2,4,5-T	9.463	9.166	3866.3E6	1694.8E6	730.573	765.149
13) T	2,4-DB	10.017	9.673	649.5E6	253.2E6	770.917	780.255
14) T	DINOSEB	11.170	10.001	2949.4E6	1261.7E6	715.254	748.402
15) T	Picloram	11.004	10.995	4740.6E6	2546.1E6	720.930	705.403
16) T	DCPA	11.435	10.927	4782.9E6	2282.8E6	736.490	760.354

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112720\
Data File : PS013344.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Nov 2020 10:01
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 11:12:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110220.M
Quant Title : 8080.M
QLast Update : Tue Nov 03 03:59:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

