

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112724\
 Data File : PS028664.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Nov 2024 18:47
 Operator : AR\AJ
 Sample : P5000-05MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-733MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/02/2024
 Supervised By :Ankita Jodhani 12/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 10:19:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 26 14:44:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	7.724	422.3E6	165.3E6	157.421	122.729
Target Compounds							
1) T	Dalapon	2.623	2.692	663.2E6	417.6E6	228.549m	191.221m
2) T	3,5-DICHL...	6.401	6.681	803.8E6	394.0E6	209.897	200.618
5) T	DICAMBA	7.415	7.923	2653.6E6	1462.3E6	231.138	237.529
6) T	MCPD	7.594	8.022	141.7E6	86893691	19.423m	20.172
7) T	MCPA	7.743	8.268	222.2E6	196.7E6	21.490	31.694 #
8) T	DICHLORPROP	8.122	8.638	600.6E6	328.3E6	192.930	205.342
9) T	2,4-D	8.352	8.969	950.4E6	604.1E6	281.850	357.271m#
10) T	Pentachlo...	8.650	9.495	4396.2E6	2347.2E6	99.705m	102.540
11) T	2,4,5-TP ...	9.229	9.877	3778.1E6	5028.0E6	206.683	526.903 #
12) T	2,4,5-T	9.521	10.291	4193.2E6	2105.8E6	223.572	226.181
13) T	2,4-DB	10.096	10.857	193.4E6	83846001	54.595	73.118m#
14) T	DINOSEB	11.304	11.237	309.8E6	169.1E6	20.617	26.941 #
15) T	Picloram	11.111	12.328	6131.9E6	2549.1E6	206.207	197.750
16) T	DCPA	11.597	12.279	6952.2E6	2946.9E6	252.851	265.550

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112724\
Data File : PS028664.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Nov 2024 18:47
Operator : AR\AJ
Sample : P5000-05MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-733MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/02/2024
Supervised By :Ankita Jodhani 12/02/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 10:19:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
Quant Title : 8080.M
QLast Update : Tue Nov 26 14:44:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

