

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112918\
 Data File : PS002488.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2018 20:21
 Operator : EI\SJ
 Sample : J6025-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 29 23:57:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 15:54:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.394	10.890	130.0E6	115.8E6	426.782	316.517 #
Target Compounds							
2) T	3,5-DICHL...	0.000	9.135	0	26783126	N.D.	59.079 #
3) T	4-Nitroph...	11.046	10.113	8888036	11624959	54.779	66.348
5) T	DICAMBA	0.000	11.259	0	4353928	N.D.	3.218 #
6) T	MCP	12.074	11.394	48124244	544186	51.276	0.600 #
7) T	MCPA	12.353	11.789	9412984	9559888	6.517	6.793
8) T	DICHLORPROP	12.999	12.404	2039051	954384	5.890	2.880 #
9) T	2,4-D	0.000	12.870	0	555057	N.D.	1.390 #
10) T	Pentachlo...	13.628	13.317	8701464	2980140	1.799	0.586 #
11) T	2,4,5-TP ...	0.000	13.957	0	3029816	N.D.	1.502 #
12) T	2,4,5-T	0.000	14.540	0	5246687	N.D.	3.356 #
14) T	DINOSEB	0.000	15.622	0	27423828	N.D.	18.323 #
15) T	Picloram	16.989	16.816	14228822	2395493	6.700	0.900 #
16) T	DCPA	17.564	16.854	23316235	1757690	10.065	0.733 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112918\
Data File : PS002488.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Nov 2018 20:21
Operator : EI\SJ
Sample : J6025-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 29 23:57:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 15:54:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T.: 9.874 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 9.135 min
Delta R.T.: 0.042 min
Response: 26783126
Conc: 59.08 ng/ml




#3 4-Nitrophenol

R.T.: 11.046 min
Delta R.T.: 0.008 min
Response: 8888036
Conc: 54.78 ng/ml



#3 4-Nitrophenol


R.T.: 10.113 min
Delta R.T.: 0.030 min
Response: 11624959
Conc: 66.35 ng/ml



#4 2,4-DCAA


R.T.: 11.394 min
Delta R.T.: -0.049 min
Response: 129997723
Conc: 426.78 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 10.890 min
Delta R.T.: -0.046 min
Response: 115817632
Conc: 316.52 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 11.752 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 11.259 min
Delta R.T.: 0.040 min
Response: 4353928
Conc: 3.22 ng/ml



#6 MCPP


R.T.: 12.074 min
Delta R.T.: -0.018 min
Response: 48124244
Conc: 51.28 ug/ml

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 11.394 min
Delta R.T.: -0.057 min
Response: 544186
Conc: 0.60 ug/ml




#7 MCPA

R.T.: 12.353 min
Delta R.T.: 0.010 min
Response: 9412984
Conc: 6.52 ug/ml



#7 MCPA


R.T.: 11.789 min
Delta R.T.: -0.037 min
Response: 9559888
Conc: 6.79 ug/ml



#8 DICHLORPROP


R.T.: 12.999 min
Delta R.T.: 0.059 min
Response: 2039051
Conc: 5.89 ng/ml

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 12.404 min
Delta R.T.: 0.061 min
Response: 954384
Conc: 2.88 ng/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 13.298 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 12.870 min
Delta R.T.: 0.056 min
Response: 555057
Conc: 1.39 ng/ml



#10 Pentachlorophenol


R.T.: 13.628 min
Delta R.T.: -0.046 min
Response: 8701464
Conc: 1.80 ng/ml

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 13.317 min
Delta R.T.: -0.042 min
Response: 2980140
Conc: 0.59 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T. : 14.543 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 13.957 min
Delta R.T.: -0.020 min
Response: 3029816
Conc: 1.50 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 14.940 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 14.540 min
Delta R.T.: 0.000 min
Response: 5246687
Conc: 3.36 ng/ml




#14 DINOSEB

R.T.: 0.000 min
Exp R.T.: 17.163 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 15.622 min
Delta R.T.: 0.004 min
Response: 27423828
Conc: 18.32 ng/ml



#15 Picloram


R.T.: 16.989 min
Delta R.T.: 0.058 min
Response: 14228822
Conc: 6.70 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 16.816 min
Delta R.T.: 0.021 min
Response: 2395493
Conc: 0.90 ng/ml



#16 DCPA

R.T.: 17.564 min
Delta R.T.: 0.059 min
Response: 23316235
Conc: 10.06 ng/ml



#16 DCPA

R.T.: 16.854 min
Delta R.T.: 0.006 min
Response: 1757690
Conc: 0.73 ng/ml