

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112918\
 Data File : PS002503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2018 04:08
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

mohammad
 11/30/2018 2:46:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 05:03:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 09 15:54:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.397	10.893	226.6E6	294.9E6	743.848	806.053
Target Compounds							
1) T	Dalapon	2.926	2.373	321.7E6	304.7E6	729.514	707.893
2) T	3,5-DICHL...	9.827	9.049	345.4E6	330.3E6	745.526m	728.473m
3) T	4-Nitroph...	10.990	10.040	123.1E6	125.6E6	758.961	716.656
5) T	DICAMBA	11.706	11.176	1087.8E6	1042.5E6	785.091	770.652m
6) T	MCPD	12.047	11.407	78595217	69038319	83.743m	76.107
7) T	MCPA	12.296	11.782	111.7E6	114.4E6	77.338m	81.316
8) T	DICHLORPROP	12.892	12.299	269.0E6	260.9E6	776.951	787.333m
9) T	2,4-D	13.250	12.769	247.4E6	303.3E6	769.767m	759.450m
10) T	Pentachlo...	13.627	13.314	3828.8E6	3932.9E6	791.536	772.804m
11) T	2,4,5-TP ...	14.496	13.933	1498.7E6	1560.2E6	797.536	773.432
12) T	2,4,5-T	14.893	14.496	1346.3E6	1249.4E6	798.693	799.252
13) T	2,4-DB	15.638	15.147	151.9E6	180.3E6	784.232m	795.472m
14) T	DINOSEB	17.114	15.574	1118.6E6	1136.5E6	783.263	759.365
15) T	Picloram	16.881	16.750	1870.2E6	2045.0E6	880.640	768.723m
16) T	DCPA	17.456	16.803	1870.2E6	2138.4E6	807.319m	891.654m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS112918\
Data File : PS002503.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2018 04:08
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
Client Sampled :
HSTDCCC750

Manual Integrations
APPROVED

mohammad
11/30/2018 2:46:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 05:03:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110918.M
Quant Title : 8080.M
QLast Update : Fri Nov 09 15:54:35 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

