

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120219\
 Data File : PS007885.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2019 17:39
 Operator : SM\AJ
 Sample : PB125203TB
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB125203TB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 02:28:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
 Quant Title : 8080.M
 QLast Update : Wed Nov 06 02:45:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.443	5.166	824.6E6	440.3E6	598.772	626.664
Target Compounds							
1) T	Dalapon	2.003	1.710	151.5E6	323.3E6	83.749	272.242 #
2) T	3,5-DICHL...	0.000	4.503	0	45792417	N.D.	50.687
5) T	DICAMBA	5.532	0.000	58192838	0	11.439	N.D. #
7) T	MCPA	5.907	5.534	37298989	29909267	6.776	9.951 #
9) T	2,4-D	0.000	6.069	0	5616557	N.D.	6.965
10) T	Pentachlo...	0.000	6.378	0	20870476	N.D.	2.131
11) T	2,4,5-TP ...	0.000	6.771	0	28706373	N.D.	7.331



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120219\
Data File : PS007885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2019 17:39
Operator : SM\AJ
Sample : PB125203TB
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB125203TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 02:28:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
Quant Title : 8080.M
QLast Update : Wed Nov 06 02:45:11 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.003 min
Delta R.T.: -0.009 min
Response: 151517536
Conc: 83.75 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB125203TB




#1 Dalapon

R.T.: 1.710 min
Delta R.T.: -0.013 min
Response: 323261453
Conc: 272.24 ng/ml


#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 4.804 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 4.503 min
Delta R.T.: 0.040 min
Response: 45792417
Conc: 50.69 ng/ml



#4 2,4-DCAA


R.T.: 5.443 min
Delta R.T.: 0.014 min
Response: 824619381
Conc: 598.77 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB125203TB




#4 2,4-DCAA

R.T.: 5.166 min
Delta R.T.: 0.015 min
Response: 440271509
Conc: 626.66 ng/ml




#5 DICAMBA

R.T.: 5.532 min
Delta R.T.: -0.046 min
Response: 58192838
Conc: 11.44 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 5.277 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 5.907 min
Delta R.T.: 0.047 min
Response: 37298989
Conc: 6.78 ug/ml

Instrument :
ECD_S
ClientSampleId :
PB125203TB




#7 MCPA

R.T.: 5.534 min
Delta R.T.: -0.002 min
Response: 29909267
Conc: 9.95 ug/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 6.414 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 6.069 min
Delta R.T.: -0.004 min
Response: 5616557
Conc: 6.96 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 6.671 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
PB125203TB




#10 Pentachlorophenol

R.T.: 6.378 min
Delta R.T.: -0.056 min
Response: 20870476
Conc: 2.13 ng/ml



#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 7.203 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)

R.T.: 6.771 min
Delta R.T.: -0.027 min
Response: 28706373
Conc: 7.33 ng/ml