

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120220\
 Data File : PS013394.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2020 18:44
 Operator : DD\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 01:11:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 01:08:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.211	6.914	923.6E6	389.5E6	1003.092	1007.589
Target Compounds							
1) T	Dalapon	3.031	2.623	981.0E6	471.6E6	916.570	908.305
2) T	3,5-DICHL...	6.427	6.067	1181.8E6	570.2E6	929.218	927.528
3) T	4-Nitroph...	7.044	6.525	418.6E6	224.7E6	923.905	929.138
5) T	DICAMBA	7.400	7.080	3445.2E6	1553.0E6	934.626	928.335
6) T	MCPD	7.569	7.176	267.1E6	126.0E6	97.305	95.287
7) T	MCPA	7.719	7.392	371.2E6	189.5E6	95.035	93.060
8) T	DICHLORPROP	8.091	7.705	861.0E6	412.3E6	941.695	937.508
9) T	2,4-D	8.323	8.002	978.9E6	504.7E6	954.934	939.604
10) T	Pentachlo...	8.630	8.427	12224.9E6	5049.9E6	913.614	903.306
11) T	2,4,5-TP ...	9.179	8.788	4979.4E6	2168.1E6	937.559	928.477
12) T	2,4,5-T	9.466	9.168	5140.1E6	2125.7E6	933.954	944.831
13) T	2,4-DB	10.019	9.676	927.7E6	285.7E6	980.360	965.359
14) T	DINOSEB	11.172	10.003	3858.8E6	1527.9E6	924.648	926.722
15) T	Picloram	11.005	10.997	7039.7E6	3375.1E6	979.837	954.898
16) T	DCPA	11.438	10.930	6329.5E6	2796.6E6	940.594	928.626

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120220\
Data File : PS013394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2020 18:44
Operator : DD\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 01:11:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 01:08:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

