

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120319\
 Data File : PS007895.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2019 10:45
 Operator : SM\AJ
 Sample : K5676-38MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SU-01-112719MSD

Manual Integrations
 APPROVED

mohammad
 12/3/2019 3:48:14 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 11:09:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
 Quant Title : 8080.M
 QLast Update : Sat Nov 23 02:27:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.430	5.152	598.1E6	288.8E6	434.305	411.094
Target Compounds							
1) T	Dalapon	2.008	1.720	816.3E6	676.7E6	451.176	569.932 #
2) T	3,5-DICHL...	4.802	4.462	1035.4E6	549.0E6	580.369	607.664
5) T	DICAMBA	5.576	5.276	2863.9E6	1364.9E6	562.948	515.780
6) T	MCP	5.725	5.358	231.9E6	113.4E6	60.852	58.388m
7) T	MCPA	5.856	5.531	313.0E6	163.5E6	56.873	54.392m
8) T	DICHLORPROP	6.199	5.804	779.0E6	372.0E6	616.357	567.302
9) T	2,4-D	6.415	6.073	943.1E6	505.3E6	672.423	626.571
10) T	Pentachlo...	6.669	6.434	9107.0E6	4285.6E6	480.553	437.642
11) T	2,4,5-TP ...	7.201	6.797	5160.4E6	2292.5E6	686.680	585.436
12) T	2,4,5-T	7.473	7.156	4822.1E6	2131.1E6	589.063	577.091
13) T	2,4-DB	8.002	7.640	562.6E6	348.8E6	526.039	713.478 #
14) T	DINOSEB	9.081	7.947	2134.3E6	899.6E6	355.400	309.456
15) T	Picloram	8.923	8.884	5215.7E6	2638.7E6	469.748	450.368
16) T	DCPA	9.353	8.837	4017.4E6	1885.0E6	409.596	360.747

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120319\
Data File : PS007895.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Dec 2019 10:45
Operator : SM\AJ
Sample : K5676-38MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SU-01-112719MSD

Manual Integrations
APPROVED

mohammad
12/3/2019 3:48:14 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 11:09:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110619.M
Quant Title : 8080.M
QLast Update : Sat Nov 23 02:27:05 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

