

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120324\
 Data File : PS028670.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2024 08:20
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 09:58:30 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 26 14:44:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.227	7.722	2082.5E6	938.2E6	776.356	696.526
Target Compounds							
1) T	Dalapon	2.631	2.687	2019.6E6	1420.7E6	695.967	650.545
2) T	3,5-DICHL...	6.400	6.679	2765.5E6	1277.8E6	722.120	650.579
3) T	4-Nitroph...	7.025	7.249	1218.2E6	607.8E6	708.681	666.310
5) T	DICAMBA	7.414	7.921	8438.3E6	4326.0E6	735.013	702.704
6) T	MCP	7.597	8.025	544.4E6	266.8E6	74.601	61.939
7) T	MCPA	7.747	8.270	750.4E6	378.3E6	72.577	60.951
8) T	DICHLORPROP	8.120	8.636	2231.3E6	1078.8E6	716.752	674.856
9) T	2,4-D	8.351	8.966	2415.6E6	1142.8E6	716.346	675.812
10) T	Pentachlo...	8.649	9.493	34176.1E6	16995.0E6	775.099	742.435
11) T	2,4,5-TP ...	9.227	9.870	13547.0E6	7002.7E6	741.091	733.841
12) T	2,4,5-T	9.519	10.289	13871.9E6	6719.9E6	739.620	721.772
13) T	2,4-DB	10.092	10.855	2565.1E6	774.2E6	724.262	675.147
14) T	DINOSEB	11.303	11.234	11072.8E6	4533.9E6	736.900	722.416
15) T	Picloram	11.111	12.325	21217.2E6	9173.7E6	713.499	711.652
16) T	DCPA	11.596	12.276	20563.3E6	8312.1E6	747.888	749.029

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120324\
Data File : PS028670.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Dec 2024 08:20
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 09:58:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
Quant Title : 8080.M
QLast Update : Tue Nov 26 14:44:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

