

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120524\
 Data File : PS028705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2024 18:42
 Operator : AR\AJ
 Sample : P5072-05MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-738MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/06/2024
 Supervised By : Ankita Jodhani 12/06/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 00:16:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 26 14:44:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.226	7.723	2054.8E6	566.6E6	766.056	420.608 #
Target Compounds						
1) T Dalapon	2.629	2.691	839.6E6	686.5E6	289.341m	314.344m
2) T 3,5-DICHL...	6.401	6.680	1605.0E6	692.7E6	419.087	352.706
5) T DICAMBA	7.415	7.922	4414.9E6	2273.3E6	384.552	369.262
6) T MCPP	7.596	8.023	227.9E6	129.7E6	31.228	30.100
7) T MCPA	7.745	8.266	372.1E6	188.1E6	35.987	30.306
8) T DICHLORPROP	8.122	8.636	1147.8E6	606.8E6	368.715	379.603m
9) T 2,4-D	8.350	8.966	2242.9E6	653.8E6	665.126m	386.626 #
10) T Pentachlo...	8.650	9.493	11698.9E6	5739.5E6	265.326	250.735
11) T 2,4,5-TP ...	9.229	9.870	7200.3E6	3992.2E6	393.893	418.360
12) T 2,4,5-T	9.521	10.290	7081.2E6	3569.5E6	377.553	383.393
13) T 2,4-DB	10.096	10.856	890.9E6	363.2E6	251.549	316.731 #
15) T Picloram	11.112	12.325	10387.6E6	4006.0E6	349.317	310.768m
16) T DCPA	11.596	12.276	8895.4E6	4784.7E6	323.526	431.165m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120524\
Data File : PS028705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2024 18:42
Operator : AR\AJ
Sample : P5072-05MS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-738MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/06/2024
Supervised By : Ankita Jodhani 12/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 00:16:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
Quant Title : 8080.M
QLast Update : Tue Nov 26 14:44:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

