

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120619\
 Data File : PS007946.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Dec 2019 02:37
 Operator : SM\AJ
 Sample : K5147-22
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 P001-WC011

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 03:43:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 05 04:44:22 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.439	5.148	793.5E6	450.5E6	507.243	580.593
Target Compounds							
2) T	3,5-DICHL...	0.000	4.483	0 55400248		N.D.	53.533
3) T	4-Nitroph...	0.000	4.825	0 12522180		N.D.	27.652
8) T	DICHLORPROP	0.000	5.814	0 8133767		N.D.	11.347
11) T	2,4,5-TP ...	0.000	6.746	0 42198824		N.D.	9.839
12) T	2,4,5-T	7.403	0.000	97411305	0	10.525	N.D. #
14) T	DINOSEB	0.000	7.978	0 24565967		N.D.	7.509
15) T	Picloram	8.909	0.000	56474076	0	4.671	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120619\
Data File : PS007946.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Dec 2019 02:37
Operator : SM\AJ
Sample : K5147-22
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
P001-WC011

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 03:43:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
Quant Title : 8080.M
QLast Update : Thu Dec 05 04:44:22 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T.: 4.801 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
P001-WC011




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 4.483 min
Delta R.T.: 0.022 min
Response: 55400248
Conc: 53.53 ng/ml




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 5.274 min
Response: 0
Conc: N.D.



#3 4-Nitrophenol


R.T.: 4.825 min
Delta R.T.: -0.015 min
Response: 12522180
Conc: 27.65 ng/ml



#4 2,4-DCAA


R.T.: 5.439 min
Delta R.T.: 0.014 min
Response: 793496857
Conc: 507.24 ng/ml

Instrument :
ECD_S
ClientSampleId :
P001-WC011




#4 2,4-DCAA

R.T.: 5.148 min
Delta R.T.: -0.004 min
Response: 450492666
Conc: 580.59 ng/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 6.197 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 5.814 min
Delta R.T.: 0.010 min
Response: 8133767
Conc: 11.35 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T. : 7.199 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
P001-WC011




#11 2,4,5-TP (SILVEX)

R.T.: 6.746 min
Delta R.T.: -0.052 min
Response: 42198824
Conc: 9.84 ng/ml




#12 2,4,5-T

R.T.: 7.403 min
Delta R.T.: -0.067 min
Response: 97411305
Conc: 10.53 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T. : 7.155 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 0.000 min
Exp R.T. : 9.078 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
P001-WC011




#14 DINOSEB

R.T.: 7.978 min
Delta R.T.: 0.031 min
Response: 24565967
Conc: 7.51 ng/ml



#15 Picloram

R.T.: 8.909 min
Delta R.T.: -0.001 min
Response: 56474076
Conc: 4.67 ng/ml



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 8.879 min
Response: 0
Conc: N.D.