

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120624\
 Data File : PS028727.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 19:20
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 12/09/2024
 Supervised By :Ankita Jodhani 12/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 22:58:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 26 14:44:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	7.720	2018.0E6	969.3E6	752.345m	719.592
Target Compounds							
1) T	Dalapon	2.627	2.685	2197.3E6	1496.5E6	757.224	685.260
2) T	3,5-DICHL...	6.398	6.678	2838.6E6	1329.6E6	741.204	676.937
3) T	4-Nitroph...	7.023	7.247	1254.6E6	630.9E6	729.840	691.673
5) T	DICAMBA	7.412	7.919	8688.4E6	4491.5E6	756.793	729.590
6) T	MCPD	7.595	8.024	557.8E6	277.4E6	76.431	64.407
7) T	MCPA	7.745	8.267	780.8E6	385.9E6	75.515	62.185
8) T	DICHLORPROP	8.118	8.634	2297.8E6	1120.4E6	738.112	700.880
9) T	2,4-D	8.348	8.963	2470.4E6	1181.3E6	732.595	698.575
10) T	Pentachlo...	8.647	9.490	35078.0E6	17580.0E6	795.553	767.991
11) T	2,4,5-TP ...	9.224	9.866	13903.0E6	7243.4E6	760.568	759.065
12) T	2,4,5-T	9.516	10.285	14234.6E6	6947.4E6	758.956	746.202
13) T	2,4-DB	10.088	10.851	2594.4E6	803.2E6	732.529	700.456
14) T	DINOSEB	11.298	11.230	11171.1E6	4561.6E6	743.442	726.828
15) T	Picloram	11.107	12.321	22194.7E6	9632.2E6	746.372	747.219
16) T	DCPA	11.592	12.272	21195.2E6	8664.5E6	770.870	780.781

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120624\
Data File : PS028727.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 19:20
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDCCC750

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 12/09/2024

Supervised By :Ankita Jodhani 12/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 22:58:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112624.M
Quant Title : 8080.M
QLast Update : Tue Nov 26 14:44:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

