

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120718\
 Data File : PS002544.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Dec 2018 00:30
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 12/10/2018 4:51:13 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 06:32:59 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 10 06:14:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.397	10.893	245.1E6	287.4E6	772.199	742.996m
Target Compounds							
1) T	Dalapon	2.925	2.365	299.6E6	293.6E6	650.909	645.070
2) T	3,5-DICHL...	9.827	9.050	335.4E6	331.1E6	701.884m	689.026
3) T	4-Nitroph...	10.986	10.039	112.9E6	122.3E6	712.380m	684.001m
5) T	DICAMBA	11.705	11.177	1021.1E6	1012.0E6	722.140m	709.845
6) T	MCPD	12.048	11.408	73617400	66992266	73.784m	71.658
7) T	MCPA	12.297	11.784	108.4E6	107.0E6	71.928m	74.440
8) T	DICHLORPROP	12.891	12.300	248.6E6	255.1E6	699.246m	713.156m
9) T	2,4-D	13.248	12.770	262.2E6	303.7E6	746.227m	716.916m
10) T	Pentachlo...	13.629	13.315	3655.8E6	3839.2E6	729.054m	729.738m
11) T	2,4,5-TP ...	14.496	13.934	1411.6E6	1505.0E6	727.356m	716.844
12) T	2,4,5-T	14.891	14.497	1354.2E6	1211.5E6	720.875	708.172
13) T	2,4-DB	15.637	15.146	168.9E6	176.1E6	746.903m	726.918m
14) T	DINOSEB	17.116	15.575	1017.5E6	1083.4E6	699.232	709.923
15) T	Picloram	16.880	16.751	1926.1E6	1962.5E6	758.935	713.337m
16) T	DCPA	17.458	16.803	1751.1E6	2168.0E6	720.591m	746.272

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120718\
Data File : PS002544.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Dec 2018 00:30
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
12/10/2018 4:51:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 06:32:59 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
Quant Title : 8080.M
QLast Update : Mon Dec 10 06:14:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

