

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120722\
 Data File : PS021376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Dec 2022 14:20
 Operator : AR\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 15:26:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:24:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.381	6.814	1452.0E6	843.5E6	503.165	491.753
Target Compounds							
1) T	Dalapon	2.481	2.556	1371.3E6	995.6E6	440.155	447.579
2) T	3,5-DICHL...	5.752	5.994	1840.4E6	1203.5E6	470.557	461.268
3) T	4-Nitroph...	6.220	6.429	798.1E6	517.1E6	448.633	444.786
5) T	DICAMBA	6.523	6.979	5436.5E6	3587.4E6	475.496	468.529
6) T	MCPD	6.677	7.076	422.2E6	265.8E6	44.704	44.601
7) T	MCPA	6.802	7.288	581.9E6	397.9E6	45.022	45.021
8) T	DICHLORPROP	7.120	7.604	1461.9E6	911.8E6	473.082	464.756
9) T	2,4-D	7.322	7.895	1780.9E6	1113.2E6	475.132	464.998
10) T	Pentachlo...	7.550	8.323	20863.7E6	12982.8E6	503.659	498.760
11) T	2,4,5-TP ...	8.079	8.685	8913.6E6	5371.1E6	490.351	481.198
12) T	2,4,5-T	8.339	9.062	9110.4E6	5186.0E6	490.777	478.815
13) T	2,4-DB	8.848	9.569	1709.3E6	657.9E6	467.447	452.993
14) T	DINOSEB	9.901	9.896	7578.3E6	3706.4E6	485.833	470.286
15) T	Picloram	9.737	10.869	13732.4E6	7796.0E6	483.524	480.123
16) T	DCPA	10.168	10.825	12544.4E6	6621.8E6	502.425	494.306

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120722\
Data File : PS021376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Dec 2022 14:20
Operator : AR\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 15:26:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:24:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

