

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120919\
 Data File : PS007951.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Dec 2019 09:39
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 09 12:09:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 05 04:44:22 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.428	5.155	1122.9E6	571.6E6	717.822	736.729
Target Compounds							
1) T	Dalapon	2.011	1.722	1387.3E6	820.1E6	680.281	690.410
2) T	3,5-DICHL...	4.805	4.465	1378.6E6	690.1E6	660.653	666.852
3) T	4-Nitroph...	5.286	4.848	455.2E6	287.3E6	598.899	634.415
5) T	DICAMBA	5.575	5.279	3976.3E6	2069.1E6	698.894	723.706
6) T	MCPD	5.726	5.363	297.5E6	153.1E6	72.625	73.814
7) T	MCPA	5.857	5.538	420.1E6	222.4E6	70.429	69.798
8) T	DICHLORPROP	6.198	5.807	957.0E6	503.0E6	672.564	701.741
9) T	2,4-D	6.414	6.077	1033.6E6	612.0E6	642.452	692.558
10) T	Pentachlo...	6.668	6.437	14858.8E6	7634.9E6	683.308	699.427
11) T	2,4,5-TP ...	7.201	6.800	5820.0E6	3055.9E6	684.008	712.519
12) T	2,4,5-T	7.473	7.159	6484.5E6	2830.7E6	700.644	695.479
13) T	2,4-DB	8.000	7.642	800.8E6	396.5E6	652.308	732.300
14) T	DINOSEB	9.078	7.949	4641.5E6	2236.7E6	666.725	683.735
15) T	Picloram	8.918	8.884	8633.9E6	4602.1E6	714.082	745.090
16) T	DCPA	9.351	8.839	7376.1E6	3996.3E6	669.893	696.398

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120919\
Data File : PS007951.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Dec 2019 09:39
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 09 12:09:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
Quant Title : 8080.M
QLast Update : Thu Dec 05 04:44:22 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

