

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120920\
 Data File : PS013425.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Dec 2020 10:34
 Operator : DD\AJ
 Sample : PB133462BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB133462BS

Manual Integrations
 APPROVED

Ankita
 12/10/2020 12:03:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 03:44:53 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 02:59:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.211	6.914	507.5E6	230.1E6	545.092	585.017
Target Compounds							
1) T	Dalapon	3.028	2.622	565.5E6	290.5E6	526.391	552.807
2) T	3,5-DICHL...	6.427	6.067	687.0E6	313.0E6	536.005	503.722m
3) T	4-Nitroph...	7.046	6.528	203.3E6	96835969	451.125	406.277m
5) T	DICAMBA	7.401	7.080	1843.3E6	854.1E6	499.152	508.222
6) T	MCPD	7.566	7.172	158.1E6	70775773	58.216	53.409
7) T	MCPA	7.715	7.386	211.1E6	123.2E6	54.014	59.238
8) T	DICHLORPROP	8.092	7.706	482.2E6	257.4E6	520.942	575.996
9) T	2,4-D	8.324	8.003	556.5E6	305.0E6	552.767	565.169
10) T	Pentachlo...	8.630	8.427	7467.4E6	3150.9E6	557.157	558.766
11) T	2,4,5-TP ...	9.178	8.787	2943.8E6	1308.3E6	552.228	557.269
12) T	2,4,5-T	9.466	9.169	2911.4E6	1236.8E6	526.584	554.365
13) T	2,4-DB	10.020	9.677	332.3E6	161.0E6	361.685	550.562 #
14) T	DINOSEB	11.173	10.004	2033.2E6	818.5E6	479.764	495.198
15) T	Picloram	11.011	11.003	2847.5E6	1462.5E6	413.398	420.032
16) T	DCPA	11.438	10.931	3378.2E6	1608.0E6	498.141	528.737

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS120920\
Data File : PS013425.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Dec 2020 10:34
Operator : DD\AJ
Sample : PB133462BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB133462BS

Manual Integrations
APPROVED

Ankita
12/10/2020 12:03:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 03:44:53 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 02:59:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

