

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121019\
 Data File : PS007989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Dec 2019 09:28
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

mohammad
 12/11/2019 12:27:05 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 10 10:55:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 05 04:44:22 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.424	5.153	1165.2E6	587.1E6	744.824	756.652
Target Compounds							
1) T	Dalapon	2.010	1.721	1363.6E6	825.4E6	668.627	694.933
2) T	3,5-DICHL...	4.800	4.462	1406.1E6	708.3E6	673.813	684.401
3) T	4-Nitroph...	5.280	4.844	478.4E6	294.7E6	629.475	650.761
5) T	DICAMBA	5.573	5.277	4088.1E6	2109.9E6	718.548	737.995
6) T	MCPD	5.725	5.361	299.1E6	156.8E6	73.007	75.624
7) T	MCPA	5.856	5.537	434.3E6	230.5E6	72.820	72.337
8) T	DICHLORPROP	6.194	5.805	1063.0E6	519.2E6	747.052m	724.355
9) T	2,4-D	6.409	6.074	1095.9E6	630.7E6	681.225	713.729
10) T	Pentachlo...	6.664	6.435	15043.5E6	7733.9E6	691.801	708.494
11) T	2,4,5-TP ...	7.197	6.798	5942.4E6	3111.6E6	698.395	725.507
12) T	2,4,5-T	7.468	7.156	6539.0E6	2943.5E6	706.529	723.198
13) T	2,4-DB	7.994	7.639	897.3E6	407.6E6	730.843m	752.912
14) T	DINOSEB	9.076	7.947	4635.5E6	2252.4E6	665.850	688.525
15) T	Picloram	8.912	8.881	8524.7E6	4448.6E6	705.056	720.231
16) T	DCPA	9.349	8.838	7482.4E6	4080.2E6	679.548	711.013

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121019\
Data File : PS007989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Dec 2019 09:28
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

mohammad
12/11/2019 12:27:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 10 10:55:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
Quant Title : 8080.M
QLast Update : Thu Dec 05 04:44:22 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

