

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121119\
 Data File : PS008021.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2019 12:26
 Operator : SM\AJ
 Sample : K6111-08MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 CL-01-120619MS

Manual Integrations
 APPROVED

mohammad
 12/12/2019 9:03:28 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 01:19:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 05 04:44:22 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.430	5.158	354.0E6	202.5E6	226.303	260.928
Target Compounds							
1) T	Dalapon	2.009	1.723	787.3E6	471.3E6	386.057	396.788
2) T	3,5-DICHL...	4.803	4.466	731.7E6	334.4E6	350.642	323.166m
5) T	DICAMBA	5.576	5.281	2354.8E6	1204.1E6	413.885	421.177
6) T	MCPD	5.724	5.363	172.1E6	88344087	42.004	42.601
7) T	MCPA	5.854	5.537	258.7E6	127.3E6	43.372	39.957m
8) T	DICHLORPROP	6.198	5.809	618.6E6	326.1E6	434.755	454.956
9) T	2,4-D	6.413	6.079	777.2E6	444.1E6	483.084	502.607
10) T	Pentachlo...	6.668	6.440	5724.4E6	2918.4E6	263.245	267.349
11) T	2,4,5-TP ...	7.200	6.801	4144.0E6	3658.4E6	487.031	853.014 #
12) T	2,4,5-T	7.472	7.161	4184.6E6	1893.9E6	452.144	465.320
13) T	2,4-DB	7.999	7.645	470.3E6	274.7E6	383.080	507.334 #
14) T	DINOSEB	9.079	7.951	1363.5E6	585.6E6	195.851	179.018
15) T	Picloram	8.919	8.887	4441.6E6	2323.1E6	367.348	376.116
16) T	DCPA	9.350	8.841	3377.4E6	1709.0E6	306.732	297.818

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121119\
Data File : PS008021.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2019 12:26
Operator : SM\AJ
Sample : K6111-08MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
CL-01-120619MS

Manual Integrations
APPROVED

mohammad
12/12/2019 9:03:28 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:19:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
Quant Title : 8080.M
QLast Update : Thu Dec 05 04:44:22 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

