

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121125\
 Data File : PS032722.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2025 12:10
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/15/2025
 Supervised By : Yogesh Patel 12/15/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 00:45:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121025.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 11 01:33:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.310	7.986	4981.0E6	2090.9E6	698.017	746.649
Target Compounds							
1) T	Dalapon	2.683	2.836	6141.7E6	2954.4E6	572.950	635.775
2) T	3,5-DICHL...	6.468	6.904	5715.0E6	2773.0E6	661.889	707.474
3) T	4-Nitroph...	7.109	7.502	2837.9E6	1351.5E6	630.234	685.095
5) T	DICAMBA	7.504	8.195	16696.9E6	8184.5E6	670.869	716.780
6) T	MCPD	7.688	8.294	1147.3E6	612.2E6	70.755	75.493
7) T	MCPA	7.841	8.548	1396.5E6	841.0E6	67.472	75.990
8) T	DICHLORPROP	8.225	8.928	4221.6E6	2062.9E6	658.446	696.087
9) T	2,4-D	8.458	9.268	5129.2E6	2407.4E6	691.553	701.637
10) T	Pentachlo...	8.776	9.822	52521.9E6	30340.3E6	661.204	732.071
11) T	2,4,5-TP ...	9.357	10.192	25510.5E6	12327.0E6	676.713	723.707
12) T	2,4,5-T	9.652	10.621	25823.8E6	12803.3E6	677.340	725.669
13) T	2,4-DB	10.236	11.193	4193.4E6	1756.6E6	662.841	714.591
14) T	DINOSEB	11.462	11.579	17722.4E6	8810.1E6	653.196	698.882
15) T	Picloram	11.266	12.704	29567.2E6	16177.6E6	619.054	647.855
16) T	DCPA	11.757	12.640	32706.8E6	14880.0E6	679.907m	719.279

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121125\
Data File : PS032722.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2025 12:10
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/15/2025
Supervised By : Yogesh Patel 12/15/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 00:45:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121025.M
Quant Title : 8080.M
QLast Update : Thu Dec 11 01:33:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

