

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121218\
 Data File : PS002581.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2018 04:18
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 12/14/2018 3:41:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 07:59:51 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 10 06:14:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.410	10.891	241.9E6	287.5E6	762.187	743.396
Target Compounds							
1) T	Dalapon	2.940	2.367	307.3E6	300.1E6	667.530	659.377
2) T	3,5-DICHL...	9.841	9.046	340.1E6	335.6E6	711.762	698.290m
3) T	4-Nitroph...	11.005	10.041	115.2E6	123.9E6	726.869	692.957m
5) T	DICAMBA	11.719	11.174	1020.7E6	1015.2E6	721.839	712.132
6) T	MCPD	12.060	11.405	79109622	67078242	79.288	71.750
7) T	MCPA	12.310	11.782	108.2E6	107.9E6	71.859	75.032
8) T	DICHLORPROP	12.905	12.297	251.8E6	250.8E6	708.066	701.373m
9) T	2,4-D	13.264	12.770	255.2E6	289.1E6	726.220	682.597m
10) T	Pentachlo...	13.641	13.311	3623.7E6	3830.5E6	722.657	728.091m
11) T	2,4,5-TP ...	14.510	13.932	1394.8E6	1496.9E6	718.708	712.982
12) T	2,4,5-T	14.907	14.503	1366.2E6	1073.1E6	727.231	627.275
13) T	2,4-DB	15.655	15.148	143.8E6	154.9E6	635.740	639.549m
14) T	DINOSEB	17.129	15.576	1092.4E6	1109.4E6	750.672	726.988m
15) T	Picloram	16.909	16.750	1665.0E6	1956.9E6	656.053	711.305m
16) T	DCPA	17.472	16.819	1758.2E6	1771.7E6	723.529	609.867

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121218\
Data File : PS002581.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2018 04:18
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
12/14/2018 3:41:32 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 07:59:51 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
Quant Title : 8080.M
QLast Update : Mon Dec 10 06:14:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

