

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\ps121222\
 Data File : PS021427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2022 17:47
 Operator : AR\AJ
 Sample : N5895-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-19MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/13/2022
 Supervised By : Ankita Jodhani 12/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 02:26:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:59:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.377	6.810	1056.1E6	576.4E6	363.471	334.318
Target Compounds							
1) T	Dalapon	2.478	2.556	832.8E6	641.9E6	263.548m	285.945m
2) T	3,5-DICHL...	5.749	5.991	1140.1E6	661.6E6	286.865	252.818
5) T	DICAMBA	6.519	6.973	4108.8E6	2588.6E6	357.056	338.944
6) T	MCP	6.672	7.068	171.9E6	127.6E6	18.410	21.680
7) T	MCPA	6.797	7.279	366.5E6	411.3E6	28.615	46.513 #
8) T	DICHLORPROP	7.117	7.599	854.0E6	464.7E6	271.648	235.942
9) T	2,4-D	7.320	7.890	1299.0E6	808.3E6	339.954	335.077
10) T	Pentachlo...	7.548	8.318	3465.8E6	2026.2E6	85.226	76.859
11) T	2,4,5-TP ...	8.077	8.680	2926.5E6	1755.7E6	158.267	156.780
12) T	2,4,5-T	8.336	9.056	4423.1E6	2453.3E6	234.427	226.346
13) T	2,4-DB	8.846	9.564	377.9E6	182.8E6	103.191m	127.197
15) T	Picloram	9.734	10.861	10499.5E6	5876.2E6	366.602	361.352
16) T	DCPA	10.167	10.819	10989.7E6	5751.4E6	431.517	423.890

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\ps121222\
Data File : PS021427.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2022 17:47
Operator : AR\AJ
Sample : N5895-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MH-19MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/13/2022
Supervised By :Ankita Jodhani 12/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 02:26:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:59:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

