

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121319\
 Data File : PS008041.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2019 15:27
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 17:43:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 05 04:44:22 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.443	5.168	1237.2E6	612.0E6	790.902	788.798
Target Compounds							
1) T	Dalapon	2.025	1.735	1523.4E6	746.2E6	747.007	628.257
2) T	3,5-DICHL...	4.819	4.477	1565.2E6	740.2E6	750.062	715.296
3) T	4-Nitroph...	5.303	4.862	460.6E6	304.8E6	605.950	673.021
5) T	DICAMBA	5.590	5.292	4430.0E6	2220.4E6	778.637	776.621
6) T	MCPD	5.741	5.376	345.8E6	172.3E6	84.394	83.089
7) T	MCPA	5.872	5.552	493.2E6	243.5E6	82.687	76.428
8) T	DICHLORPROP	6.213	5.821	1140.6E6	549.4E6	801.644	766.385
9) T	2,4-D	6.431	6.092	1181.0E6	643.5E6	734.075	728.284
10) T	Pentachlo...	6.682	6.451	16940.7E6	8343.2E6	779.047	764.311
11) T	2,4,5-TP ...	7.215	6.814	6551.2E6	3298.3E6	769.949	769.058
12) T	2,4,5-T	7.489	7.172	7241.4E6	2949.3E6	782.424	724.643
13) T	2,4-DB	8.016	7.657	782.7E6	398.8E6	637.521	736.649
14) T	DINOSEB	9.093	7.962	5212.4E6	2378.2E6	748.730	726.971
15) T	Picloram	8.933	8.899	9676.9E6	5109.8E6	800.351	827.293
16) T	DCPA	9.366	8.853	8321.7E6	4332.2E6	755.773	754.934

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121319\
Data File : PS008041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2019 15:27
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 17:43:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120419.M
Quant Title : 8080.M
QLast Update : Thu Dec 05 04:44:22 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

