

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121420\
 Data File : PS013522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2020 02:08
 Operator : DD\AJ
 Sample : PB133601BS
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB133601BS

Manual Integrations
 APPROVED

Ankita
 12/15/2020 1:06:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 04:00:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 02:59:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.210	6.913	505.3E6	230.8E6	542.687	586.697
Target Compounds							
1) T	Dalapon	3.029	2.622	546.1E6	322.0E6	508.338	612.719
2) T	3,5-DICHL...	6.426	6.066	674.8E6	305.0E6	526.467	490.787m
3) T	4-Nitroph...	7.045	6.526	214.0E6	98934487	474.912	415.082
5) T	DICAMBA	7.400	7.079	1804.0E6	835.7E6	488.513	497.246
6) T	MCPD	7.566	7.171	159.9E6	68941080	58.871	52.024
7) T	MCPA	7.714	7.385	204.8E6	125.3E6	52.397	60.272
8) T	DICHLORPROP	8.091	7.704	463.4E6	227.1E6	500.679	508.194
9) T	2,4-D	8.323	8.001	544.4E6	294.8E6	540.829	546.363
10) T	Pentachlo...	8.629	8.425	7279.4E6	3105.8E6	543.135	550.772
11) T	2,4,5-TP ...	9.178	8.786	2843.0E6	1273.8E6	533.313	542.570
12) T	2,4,5-T	9.465	9.168	2820.0E6	1205.9E6	510.047	540.500
13) T	2,4-DB	10.018	9.675	327.3E6	155.3E6	356.246	531.055 #
14) T	DINOSEB	11.171	10.003	2021.7E6	826.3E6	477.053	499.894
15) T	Picloram	11.008	11.001	2759.3E6	1445.6E6	400.604	415.197
16) T	DCPA	11.437	10.929	3306.4E6	1595.5E6	487.557	524.631

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121420\
Data File : PS013522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2020 02:08
Operator : DD\AJ
Sample : PB133601BS
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB133601BS

Manual Integrations
APPROVED

Ankita
12/15/2020 1:06:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 04:00:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 02:59:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

