

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121421\
 Data File : PS017672.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2021 12:20
 Operator : AJ\MA
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 13:17:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
 Quant Title : 8080.M
 QLast Update : Tue Dec 14 13:15:58 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.130	6.717	1550.8E6	606.1E6	1005.244	1026.358
Target Compounds							
1) T	Dalapon	2.978	2.486	2325.1E6	721.1E6	915.479	918.989
2) T	3,5-DICHL...	6.357	5.920	2051.7E6	855.4E6	928.882	943.010
3) T	4-Nitroph...	6.961	6.346	638.4E6	320.8E6	941.458	925.048
5) T	DICAMBA	7.317	6.876	5910.0E6	2506.7E6	934.261	936.987
6) T	MCPD	7.488	6.974	463.3E6	209.1E6	97.656	96.446
7) T	MCPA	7.637	7.186	627.0E6	289.8E6	95.262	95.189
8) T	DICHLORPROP	8.007	7.491	1454.7E6	644.5E6	944.983	956.358
9) T	2,4-D	8.237	7.782	1654.3E6	774.8E6	947.206	954.842
10) T	Pentachlo...	8.538	8.202	19595.3E6	7700.9E6	912.028	904.112
11) T	2,4,5-TP ...	9.088	8.564	7968.0E6	3432.0E6	938.807	939.348
12) T	2,4,5-T	9.373	8.941	8182.0E6	3282.7E6	937.533	947.363
13) T	2,4-DB	9.922	9.444	1489.5E6	440.2E6	998.391	977.346
14) T	DINOSEB	11.068	9.767	6282.2E6	2437.0E6	930.492	939.836
15) T	Picloram	10.894	10.746	10769.7E6	4813.3E6	980.960	964.668
16) T	DCPA	11.336	10.694	10139.9E6	4204.3E6	940.496	937.312

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121421\
Data File : PS017672.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2021 12:20
Operator : AJ\MA
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 13:17:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
Quant Title : 8080.M
QLast Update : Tue Dec 14 13:15:58 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

