

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121421\
 Data File : PS017673.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2021 12:44
 Operator : AJ\MA
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 13:17:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
 Quant Title : 8080.M
 QLast Update : Tue Dec 14 13:15:58 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.128	6.716	2289.5E6	915.6E6	1484.079	1550.592
Target Compounds							
1) T	Dalapon	2.978	2.484	3486.2E6	1082.9E6	1372.635	1380.150
2) T	3,5-DICHL...	6.356	5.919	2991.4E6	1266.1E6	1354.282	1395.832
3) T	4-Nitroph...	6.958	6.344	994.3E6	478.1E6	1466.430	1378.822
5) T	DICAMBA	7.316	6.875	8585.9E6	3613.0E6	1357.274	1350.501
6) T	MCP	7.492	6.978	712.5E6	318.6E6	150.178	146.948
7) T	MCPA	7.641	7.191	948.7E6	438.5E6	144.152	144.031
8) T	DICHLORPROP	8.006	7.490	2155.3E6	963.1E6	1400.068	1429.112
9) T	2,4-D	8.234	7.780	2454.9E6	1159.4E6	1405.619	1428.799
10) T	Pentachlo...	8.538	8.202	27047.6E6	10417.2E6	1258.883	1223.012
11) T	2,4,5-TP ...	9.086	8.563	11446.3E6	4898.9E6	1348.624	1340.835
12) T	2,4,5-T	9.371	8.939	11721.3E6	4707.7E6	1343.079	1358.612
13) T	2,4-DB	9.918	9.442	2289.4E6	669.0E6	1534.554	1485.401
14) T	DINOSEB	11.067	9.766	9060.1E6	3539.6E6	1341.941	1365.057
15) T	Picloram	10.891	10.743	15886.3E6	6888.0E6	1447.005	1380.486
16) T	DCPA	11.335	10.693	14395.5E6	5934.7E6	1335.207	1323.089

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121421\
Data File : PS017673.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2021 12:44
Operator : AJ\MA
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 13:17:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
Quant Title : 8080.M
QLast Update : Tue Dec 14 13:15:58 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

