

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121421\
 Data File : PS017676.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2021 13:57
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 15:43:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
 Quant Title : 8080.M
 QLast Update : Tue Dec 14 13:29:21 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.130	6.717	1158.1E6	445.0E6	740.954	746.599
Target Compounds							
1) T	Dalapon	2.978	2.484	1727.4E6	533.8E6	674.099	689.601
2) T	3,5-DICHL...	6.358	5.920	1540.4E6	634.1E6	689.463	699.419
3) T	4-Nitroph...	6.965	6.347	456.2E6	236.6E6	674.628	679.542
5) T	DICAMBA	7.317	6.876	4456.6E6	1892.8E6	704.775	710.507
6) T	MCPD	7.486	6.971	334.5E6	153.3E6	70.951	70.427
7) T	MCPA	7.633	7.182	459.5E6	212.9E6	69.473	69.627
8) T	DICHLORPROP	8.007	7.491	1084.9E6	476.4E6	694.313	698.996
9) T	2,4-D	8.238	7.783	1226.6E6	572.1E6	693.812	703.629
10) T	Pentachlo...	8.537	8.202	15306.5E6	6078.1E6	709.380	709.410
11) T	2,4,5-TP ...	9.088	8.563	6040.1E6	2607.7E6	710.022	717.319
12) T	2,4,5-T	9.373	8.940	6196.9E6	2465.3E6	712.343	720.909
13) T	2,4-DB	9.923	9.444	1049.5E6	320.8E6	710.686	708.164
14) T	DINOSEB	11.067	9.766	4737.1E6	1819.2E6	692.772	704.157
15) T	Picloram	10.894	10.745	7810.5E6	3570.3E6	719.780	726.940
16) T	DCPA	11.335	10.693	7741.7E6	3225.2E6	712.827	720.089

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121421\
Data File : PS017676.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2021 13:57
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 15:43:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121421.M
Quant Title : 8080.M
QLast Update : Tue Dec 14 13:29:21 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

