

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121523\
 Data File : PS025128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Dec 2023 21:44
 Operator : AR\AJ
 Sample : 05803-03MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 600572408-MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/17/2023
 Supervised By : Ankita Jodhani 12/18/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 22:33:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112823.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 28 16:40:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.356	6.768	1613.9E6	296.2E6	721.183m	408.809m#
Target Compounds							
1) T	Dalapon	2.473	2.530	795.8E6	490.7E6	275.948m	427.763m#
2) T	3,5-DICHL...	5.729	5.958	1606.0E6	560.7E6	475.146	493.279
5) T	DICAMBA	6.495	6.929	4041.9E6	1290.7E6	414.742	416.572
6) T	MCP	6.645	7.027	374.2E6	102.4E6	54.639m	49.354
7) T	MCPA	6.768	7.229	352.6E6	149.1E6	37.033m	45.316
8) T	DICHLORPROP	7.081	7.542	1296.8E6	404.6E6	505.532	514.387
9) T	2,4-D	7.278	7.823	1449.3E6	522.2E6	477.171	536.104
10) T	Pentachlo...	7.507	8.246	17341.8E6	5296.8E6	494.584	467.276
11) T	2,4,5-TP ...	8.029	8.605	8116.2E6	2261.2E6	534.276	512.725m
12) T	2,4,5-T	8.291	8.976	6972.9E6	2051.3E6	482.743	475.445
13) T	2,4-DB	8.796	9.481	1469.3E6	703.0E6	632.419m	1270.092 #
14) T	DINOSEB	9.841	9.825	3752.4E6	5226.5E6	345.597	1704.715 #
15) T	Picloram	9.668	10.755	8633.4E6	4269.6E6	443.062m	652.321 #
16) T	DCPA	10.107	10.725	9749.7E6	2589.3E6	519.623	477.732

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121523\
Data File : PS025128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2023 21:44
Operator : AR\AJ
Sample : 05803-03MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
600572408-MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/17/2023
Supervised By :Ankita Jodhani 12/18/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 22:33:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112823.M
Quant Title : 8080.M
QLast Update : Tue Nov 28 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

