

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121622\
 Data File : PS021448.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Dec 2022 09:34
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 16 20:33:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:59:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.378	6.810	2109.6E6	1313.9E6	726.066	762.147
Target Compounds							
1) T	Dalapon	2.479	2.555	2047.7E6	1460.1E6	647.977	650.405
2) T	3,5-DICHL...	5.749	5.991	2832.4E6	1882.3E6	712.655	719.312
3) T	4-Nitroph...	6.217	6.425	1203.6E6	809.8E6	669.388	694.929
5) T	DICAMBA	6.519	6.974	8513.3E6	5627.8E6	739.819	736.900
6) T	MCPD	6.678	7.074	671.1E6	437.1E6	71.879	74.293
7) T	MCPA	6.803	7.287	929.3E6	613.5E6	72.550	69.380
8) T	DICHLORPROP	7.117	7.599	2258.6E6	1410.5E6	718.451	716.105
9) T	2,4-D	7.320	7.891	2708.1E6	1740.0E6	708.704	721.312
10) T	Pentachlo...	7.548	8.318	30667.6E6	19217.2E6	754.139	728.969
11) T	2,4,5-TP ...	8.077	8.680	13505.8E6	8219.6E6	730.416	733.989
12) T	2,4,5-T	8.337	9.056	13854.3E6	7927.3E6	734.294	731.399
13) T	2,4-DB	8.846	9.563	2516.1E6	1087.5E6	687.030	756.548
14) T	DINOSEB	9.900	9.890	11568.6E6	5777.9E6	726.088	730.530
15) T	Picloram	9.735	10.862	20300.1E6	11573.2E6	708.804	711.685
16) T	DCPA	10.168	10.818	18506.2E6	9878.3E6	726.657	728.048

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121622\
Data File : PS021448.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Dec 2022 09:34
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 20:33:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:59:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

