

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121718\
 Data File : PS002631.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2018 19:39
 Operator : EI\SJ
 Sample : J6347-15MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 P001-TW001-01MS

Manual Integrations
 APPROVED

Sohil
 12/18/2018 4:52:36 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 02:19:30 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 10 06:14:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.396	10.893	173.1E6	185.4E6	545.472	479.375
Target Compounds							
1) T	Dalapon	2.925	2.372	181.3E6	199.3E6	393.919	437.911
2) T	3,5-DICHL...	9.824	9.048	233.4E6	229.2E6	488.401m	477.049
3) T	4-Nitroph...	10.993	10.043	73253718	74405590	462.347	416.284
5) T	DICAMBA	11.704	11.175	650.1E6	679.6E6	459.778	476.733m
6) T	MCPD	12.077	11.405	55639178	47854927	55.765m	51.188m
7) T	MCPA	12.292	11.779	65254738	74918594	43.319m	52.119
8) T	DICHLORPROP	12.890	12.299	168.9E6	158.1E6	474.905	441.948
9) T	2,4-D	13.251	12.771	180.8E6	205.3E6	514.708	484.682
10) T	Pentachlo...	13.626	13.314	2544.1E6	2776.3E6	507.354	527.714
11) T	2,4,5-TP ...	14.495	13.934	921.1E6	1081.2E6	474.602	514.955
12) T	2,4,5-T	14.894	14.508	886.6E6	670.7E6	471.960	392.019
13) T	2,4-DB	15.643	15.154	89226039	89101164	394.519	367.804
14) T	DINOSB	17.114	15.577	743.3E6	715.8E6	510.813	469.063
15) T	Picloram	16.912	16.753	811.8E6	907.2E6	319.888	329.742
16) T	DCPA	17.457	16.831	708.8E6	953.3E6	291.675	328.132

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121718\
Data File : PS002631.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2018 19:39
Operator : EI\SJ
Sample : J6347-15MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
P001-TW001-01MS

Manual Integrations
APPROVED

Sohil
12/18/2018 4:52:36 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 02:19:30 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
Quant Title : 8080.M
QLast Update : Mon Dec 10 06:14:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

