

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121818\
 Data File : PS002654.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Dec 2018 21:52
 Operator : EI\SJ
 Sample : J6374-03MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FK-GF-02-029-BMS

Manual Integrations
 APPROVED

Sohil
 12/19/2018 10:40:53 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 04:19:23 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 10 06:14:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.396	10.893	33939414	40647859	106.923m	105.096
Target Compounds							
1) T	Dalapon	2.920	2.356	43537319	85866136	94.581	188.663m#
2) T	3,5-DICHL...	9.827	9.050	32464455	30743769	67.945	63.976m
3) T	4-Nitroph...	11.001	10.048	8775250	18891969	55.386	105.697 #
5) T	DICAMBA	11.702	11.176	85958989	94041768	60.792m	65.966
6) T	MCPD	12.076	11.451	16890759	1789355	16.929m	1.914m#
7) T	MCPA	12.289	11.829	6734364	8923668	4.471m	6.208m#
8) T	DICHLORPROP	12.890	12.301	25384894	26294007	71.389m	73.518
9) T	2,4-D	13.256	12.773	19268215	27467805	54.840	64.849m
10) T	Pentachlo...	13.625	13.314	528.9E6	567.1E6	105.472	107.800
11) T	2,4,5-TP ...	14.495	13.934	145.2E6	165.5E6	74.836	78.809
12) T	2,4,5-T	14.896	14.507	105.3E6	160.9E6	56.075	94.077m#
13) T	2,4-DB	15.647	15.151	11787434	11833890	52.119	48.850
14) T	DINOSEB	17.113	15.575	29906233	28751948	20.551	18.841
15) T	Picloram	16.896	16.752	179.0E6	202.8E6	70.552	73.723
16) T	DCPA	17.456	16.816	132.8E6	244.7E6	54.650	84.220m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121818\
Data File : PS002654.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Dec 2018 21:52
Operator : EI\SJ
Sample : J6374-03MS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
FK-GF-02-029-BMS

Manual Integrations
APPROVED

Sohil
12/19/2018 10:40:53 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 04:19:23 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
Quant Title : 8080.M
QLast Update : Mon Dec 10 06:14:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

