

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122018\
 Data File : PS002694.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2018 03:51
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 12/21/2018 2:42:22 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 09:17:00 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 10 06:14:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.422	10.920	247.9E6	319.3E6	781.104	825.556
Target Compounds							
1) T	Dalapon	2.951	2.398	338.3E6	336.2E6	735.031	738.630
2) T	3,5-DICHL...	9.852	9.076	353.7E6	365.8E6	740.197	761.271
3) T	4-Nitroph...	11.022	10.072	98592403	123.4E6	622.275	690.157
5) T	DICAMBA	11.729	11.203	1091.0E6	1134.6E6	771.575	795.843
6) T	MCPD	12.069	11.433	75661884	74027919	75.833	79.184
7) T	MCPA	12.319	11.809	110.7E6	106.6E6	73.464	74.132
8) T	DICHLORPROP	12.915	12.326	263.8E6	276.6E6	741.781	773.250
9) T	2,4-D	13.277	12.799	249.4E6	303.2E6	709.955	715.813
10) T	Pentachlo...	13.651	13.341	3881.9E6	4185.0E6	774.147	795.478
11) T	2,4,5-TP ...	14.521	13.961	1454.9E6	1644.5E6	749.655	783.247
12) T	2,4,5-T	14.922	14.536	1374.7E6	1340.2E6	731.797	783.382m
13) T	2,4-DB	15.674	15.185	148.7E6	182.4E6	657.656m	753.128m
14) T	DINOSEB	17.139	15.604	1077.0E6	1178.1E6	740.073m	771.967
15) T	Picloram	16.948	16.780	1400.3E6	2128.1E6	551.750	773.513 #
16) T	DCPA	17.483	16.864	1867.3E6	2001.9E6	768.410	689.099m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122018\
Data File : PS002694.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2018 03:51
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
12/21/2018 2:42:22 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 09:17:00 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120618.M
Quant Title : 8080.M
QLast Update : Mon Dec 10 06:14:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

