

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122022\
 Data File : PS021503.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2022 00:50
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 03:36:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:59:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.376	6.809	2180.2E6	1397.7E6	750.376	810.738
Target Compounds							
1) T	Dalapon	2.479	2.555	2224.8E6	1609.9E6	704.030	717.134
2) T	3,5-DICHL...	5.747	5.990	2931.0E6	2004.9E6	737.480	766.177
3) T	4-Nitroph...	6.215	6.424	1231.3E6	844.8E6	684.766	724.919
5) T	DICAMBA	6.517	6.973	8825.3E6	5961.5E6	766.929	780.584
6) T	MCPD	6.676	7.074	689.5E6	463.6E6	73.848	78.785
7) T	MCPA	6.802	7.286	938.5E6	643.9E6	73.263	72.820
8) T	DICHLORPROP	7.115	7.598	2317.2E6	1494.8E6	737.086	758.936
9) T	2,4-D	7.317	7.889	2772.2E6	1827.1E6	725.491	757.410
10) T	Pentachlo...	7.545	8.316	31644.8E6	20305.0E6	778.168	770.231
11) T	2,4,5-TP ...	8.073	8.678	13897.2E6	8685.0E6	751.579	775.549
12) T	2,4,5-T	8.333	9.054	14225.2E6	8365.0E6	753.952	771.781
13) T	2,4-DB	8.841	9.561	2612.8E6	1147.0E6	713.431	797.945
14) T	DINOSEB	9.896	9.888	11852.7E6	6075.8E6	743.920	768.194
15) T	Picloram	9.730	10.858	19149.0E6	11181.1E6	668.611	687.573
16) T	DCPA	10.162	10.816	19153.1E6	10490.0E6	752.058	773.132

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122022\
Data File : PS021503.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2022 00:50
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 03:36:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:59:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

