

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122123\
 Data File : PS025150.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Dec 2023 14:15
 Operator : AR\AJ
 Sample : 05887-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-121423MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 12/22/2023
 Supervised By : Ankita Jodhani 12/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 22:32:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112823.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 28 16:40:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.353	6.767	1128.4E6	323.2E6	504.230m	446.121
Target Compounds							
1) T	Dalapon	2.464	2.518	1510.4E6	761.8E6	523.768m	664.115m#
2) T	3,5-DICHL...	5.729	5.958	1255.0E6	387.1E6	371.317	340.556
5) T	DICAMBA	6.494	6.928	3624.0E6	1104.8E6	371.864m	356.580
6) T	MCP	6.644	7.023	205.1E6	60362502	29.939	29.101
7) T	MCPA	6.766	7.228	338.9E6	111.3E6	35.600	33.843
8) T	DICHLORPROP	7.081	7.541	848.5E6	274.1E6	330.758	348.492
9) T	2,4-D	7.278	7.823	1180.3E6	383.7E6	388.584	393.923
10) T	Pentachlo...	7.507	8.245	6923.6E6	2027.1E6	197.459	178.831
11) T	2,4,5-TP ...	8.028	8.604	4206.6E6	1219.0E6	276.913m	276.421
12) T	2,4,5-T	8.284	8.973	5027.4E6	1412.3E6	348.051	327.326
13) T	2,4-DB	8.790	9.476	615.2E6	156.4E6	264.810	282.504
15) T	Picloram	9.667	10.752	7055.4E6	2237.7E6	362.081	341.882
16) T	DCPA	10.105	10.724	8356.8E6	2279.8E6	445.385m	420.637

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122123\
Data File : PS025150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Dec 2023 14:15
Operator : AR\AJ
Sample : 05887-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WC-121423MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 12/22/2023
Supervised By :Ankita Jodhani 12/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 22:32:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112823.M
Quant Title : 8080.M
QLast Update : Tue Nov 28 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

