

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122220\
 Data File : PS013575.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Dec 2020 13:34
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 22 13:51:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 02:59:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.210	6.912	732.1E6	310.0E6	786.279	788.091
Target Compounds							
1) T	Dalapon	3.032	2.625	749.3E6	363.8E6	697.511	692.174
2) T	3,5-DICHL...	6.427	6.067	943.0E6	456.6E6	735.731	734.837
3) T	4-Nitroph...	7.045	6.526	321.9E6	170.0E6	714.534	713.414
5) T	DICAMBA	7.400	7.078	2779.5E6	1262.4E6	752.673	751.162
6) T	MCPD	7.567	7.173	212.4E6	100.6E6	78.205	75.943
7) T	MCPA	7.716	7.388	299.2E6	151.7E6	76.547	72.935
8) T	DICHLORPROP	8.091	7.704	690.9E6	334.7E6	746.465	749.015
9) T	2,4-D	8.323	8.002	771.9E6	405.5E6	766.793	751.566
10) T	Pentachlo...	8.629	8.425	10160.3E6	4258.0E6	758.087	755.089
11) T	2,4,5-TP ...	9.178	8.787	4018.6E6	1782.0E6	753.851	759.032
12) T	2,4,5-T	9.466	9.168	4139.0E6	1718.1E6	748.610	770.079
13) T	2,4-DB	10.021	9.676	648.5E6	226.4E6	705.969	774.101
14) T	DINOSEB	11.171	10.002	3165.2E6	1278.3E6	746.879	773.400
15) T	Picloram	11.008	10.999	4967.8E6	2475.5E6	721.235	710.989
16) T	DCPA	11.437	10.928	5114.6E6	2331.5E6	754.188	766.634

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122220\
Data File : PS013575.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Dec 2020 13:34
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 22 13:51:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 02:59:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

