

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122320\
 Data File : PS013601.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Dec 2020 00:28
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 00:53:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 02:59:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.209	6.912	730.7E6	306.5E6	784.802	779.301
Target Compounds							
1) T	Dalapon	3.031	2.622	681.2E6	331.3E6	634.126	630.397
2) T	3,5-DICHL...	6.426	6.066	940.2E6	450.7E6	733.542	725.268
3) T	4-Nitroph...	7.044	6.526	320.5E6	170.0E6	711.345	713.253
5) T	DICAMBA	7.399	7.077	2779.7E6	1245.2E6	752.707	740.934
6) T	MCPD	7.566	7.172	210.9E6	98155519	77.661	74.070
7) T	MCPA	7.715	7.387	298.4E6	146.5E6	76.349	70.458
8) T	DICHLORPROP	8.090	7.703	688.4E6	328.2E6	743.752	734.284
9) T	2,4-D	8.322	8.001	751.3E6	396.6E6	746.305	735.033
10) T	Pentachlo...	8.628	8.424	10127.1E6	4209.1E6	755.605	746.414
11) T	2,4,5-TP ...	9.177	8.785	4011.5E6	1753.5E6	752.509	746.886
12) T	2,4,5-T	9.464	9.167	4081.8E6	1682.4E6	738.270	754.065
13) T	2,4-DB	10.019	9.674	632.2E6	223.7E6	688.254	764.914
14) T	DINOSEB	11.169	10.001	3095.8E6	1230.5E6	730.516	744.422
15) T	Picloram	11.007	10.998	4521.5E6	2228.0E6	656.441	639.915
16) T	DCPA	11.435	10.928	5125.3E6	2302.6E6	755.758	757.134

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122320\
Data File : PS013601.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Dec 2020 00:28
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 00:53:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 02:59:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

