

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122322\
 Data File : PS021543.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2022 11:00
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 01:16:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 07 15:59:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.375	6.808	2060.1E6	1319.5E6	709.022	765.353
Target Compounds							
1) T	Dalapon	2.479	2.556	2122.9E6	1533.8E6	671.796	683.246
2) T	3,5-DICHL...	5.747	5.989	2776.0E6	1893.0E6	698.475	723.419
3) T	4-Nitroph...	6.214	6.424	1162.9E6	803.6E6	646.714	689.619
5) T	DICAMBA	6.517	6.972	8325.6E6	5633.4E6	723.506	737.633
6) T	MCPD	6.675	7.072	639.1E6	433.9E6	68.446	73.737
7) T	MCPA	6.801	7.285	881.0E6	605.4E6	68.780	68.468
8) T	DICHLORPROP	7.115	7.598	2210.8E6	1426.6E6	703.230	724.321
9) T	2,4-D	7.317	7.889	2654.4E6	1747.5E6	694.654	724.392
10) T	Pentachlo...	7.546	8.317	30173.7E6	19368.4E6	741.992	734.702
11) T	2,4,5-TP ...	8.074	8.679	13235.9E6	8284.3E6	715.819	739.768
12) T	2,4,5-T	8.334	9.055	13529.2E6	7980.2E6	717.060	736.274
13) T	2,4-DB	8.843	9.562	2512.9E6	1072.4E6	686.174	746.042
14) T	DINOSEB	9.898	9.890	11204.5E6	5787.7E6	703.237	731.770
15) T	Picloram	9.732	10.861	18402.9E6	10610.3E6	642.558	652.474
16) T	DCPA	10.165	10.818	18195.0E6	9990.2E6	714.437	736.294

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122322\
Data File : PS021543.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2022 11:00
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 01:16:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120722.M
Quant Title : 8080.M
QLast Update : Wed Dec 07 15:59:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

