

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122324\
 Data File : PS028789.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2024 12:11
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 23 13:35:57 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122324.M
 Quant Title : 8080.M
 QLast Update : Mon Dec 23 13:35:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.210	7.700	1691.0E6	842.4E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.620	2.674	1942.8E6	1349.3E6	682.500	682.500
2) T	3,5-DICHL...	6.385	6.660	2285.6E6	1155.6E6	697.500	697.500
3) T	4-Nitroph...	7.009	7.228	1007.9E6	559.5E6	682.500	682.500
5) T	DICAMBA	7.397	7.898	7011.9E6	3936.4E6	705.000	705.000
6) T	MCPD	7.579	8.002	455.0E6	235.2E6	70.500	70.500
7) T	MCPA	7.729	8.245	625.6E6	327.4E6	69.750	69.750
8) T	DICHLORPROP	8.102	8.612	1852.8E6	978.9E6	705.000	705.000
9) T	2,4-D	8.332	8.940	2003.9E6	1024.2E6	705.000	705.000
10) T	Pentachlo...	8.629	9.465	28730.9E6	15836.1E6	712.500	712.500
11) T	2,4,5-TP ...	9.206	9.842	11258.5E6	6490.8E6	712.500	712.500
12) T	2,4,5-T	9.497	10.260	11647.7E6	6266.0E6	712.500	712.500
13) T	2,4-DB	10.069	10.826	2161.1E6	705.6E6	712.500	712.500
14) T	DINOSEB	11.276	11.204	9500.8E6	4334.9E6	705.000	705.000
15) T	Picloram	11.085	12.292	19296.4E6	9378.2E6	712.500	712.500
16) T	DCPA	11.571	12.245	17447.7E6	7888.2E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122324\
Data File : PS028789.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2024 12:11
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 23 13:35:57 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS122324.M
Quant Title : 8080.M
QLast Update : Mon Dec 23 13:35:06 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

