

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122420\
 Data File : PS013608.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Dec 2020 17:10
 Operator : DD\AJ
 Sample : L5189-06
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 17:49:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
 Quant Title : 8080.M
 QLast Update : Thu Dec 03 02:59:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.210	6.913	224.3E6	106.4E6	240.867	270.470
Target Compounds							
1) T	Dalapon	2.989f	2.585	68641577	33946964	63.900	64.596
2) T	3,5-DICHL...	6.469f	6.095	10399855	34009191	8.114	54.733 #
3) T	4-Nitroph...	6.983f	0.000	5361872	0	11.900	N.D. #
5) T	DICAMBA	0.000	7.046	0	6062144	N.D.	3.607
6) T	MCPD	7.563	0.000	8897407	0	3.277	N.D. #
7) T	MCPA	7.691	7.370	12068567	20563469	3.088	9.888 #
8) T	DICHLORPROP	8.104	7.708	6776269	3609516	7.321	8.077
10) T	Pentachlo...	8.632	8.425	11927020	3915094	<MDL	<MDL #
11) T	2,4,5-TP ...	0.000	8.761	0	9824126	N.D.	4.185
12) T	2,4,5-T	9.398f	0.000	23647104	0	4.277	N.D. #
14) T	DINOSEB	11.149	10.029	31120949	19815410	7.344	11.988 #
15) T	Picloram	11.027	11.016	52079405	3813008	7.561	1.095 #
16) T	DCPA	0.000	10.933	0	5471999	N.D.	1.799



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122420\
Data File : PS013608.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Dec 2020 17:10
Operator : DD\AJ
Sample : L5189-06
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 17:49:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS120220.M
Quant Title : 8080.M
QLast Update : Thu Dec 03 02:59:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.989 min
Delta R.T.: -0.042 min
Response: 68641577
Conc: 63.90 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.585 min
Delta R.T.: -0.037 min
Response: 33946964
Conc: 64.60 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.469 min
Delta R.T.: 0.042 min
Response: 10399855
Conc: 8.11 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.095 min
Delta R.T.: 0.027 min
Response: 34009191
Conc: 54.73 ng/ml



#3 4-Nitrophenol


R.T.: 6.983 min
Delta R.T.: -0.063 min
Response: 5361872
Conc: 11.90 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 6.527 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.210 min
Delta R.T.: 0.000 min
Response: 224261488
Conc: 240.87 ng/ml



#4 2,4-DCAA


R.T.: 6.913 min
Delta R.T.: 0.000 min
Response: 106392602
Conc: 270.47 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 7.401 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.046 min
Delta R.T.: -0.034 min
Response: 6062144
Conc: 3.61 ng/ml




#6 MCP

R.T.: 7.563 min
Delta R.T.: -0.005 min
Response: 8897407
Conc: 3.28 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.174 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.691 min
Delta R.T.: -0.025 min
Response: 12068567
Conc: 3.09 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.370 min
Delta R.T.: -0.019 min
Response: 20563469
Conc: 9.89 ug/ml




#8 DICHLORPROP

R.T.: 8.104 min
Delta R.T.: 0.013 min
Response: 6776269
Conc: 7.32 ng/ml



#8 DICHLORPROP


R.T.: 7.708 min
Delta R.T.: 0.002 min
Response: 3609516
Conc: 8.08 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 9.179 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 8.761 min
Delta R.T.: -0.028 min
Response: 9824126
Conc: 4.18 ng/ml




#12 2,4,5-T

R.T.: 9.398 min
Delta R.T.: -0.069 min
Response: 23647104
Conc: 4.28 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 9.170 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 11.149 min
Delta R.T.: -0.022 min
Response: 31120949
Conc: 7.34 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 10.029 min
Delta R.T.: 0.025 min
Response: 19815410
Conc: 11.99 ng/ml




#15 Picloram

R.T.: 11.027 min
Delta R.T.: 0.018 min
Response: 52079405
Conc: 7.56 ng/ml



#15 Picloram


R.T.: 11.016 min
Delta R.T.: 0.015 min
Response: 3813008
Conc: 1.10 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 11.437 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 10.933 min
Delta R.T.: 0.003 min
Response: 5471999
Conc: 1.80 ng/ml