

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122523\
 Data File : PS025208.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2023 02:21
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 12/26/2023
 Supervised By :Ankita Jodhani 12/26/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 26 03:36:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112823.M
 Quant Title : 8080.M
 QLast Update : Tue Nov 28 16:40:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.353	6.766	1637.8E6	526.7E6	731.846	726.967
Target Compounds							
1) T	Dalapon	2.472	2.530	1955.8E6	751.9E6	678.219	655.473
2) T	3,5-DICHL...	5.728	5.956	2309.3E6	785.8E6	683.228	691.261
3) T	4-Nitroph...	6.191	6.382	1045.3E6	321.5E6	709.813	660.090
5) T	DICAMBA	6.494	6.925	6848.6E6	2155.3E6	702.741	695.651m
6) T	MCPD	6.645	7.023	479.6E6	144.1E6	70.017	69.454
7) T	MCPA	6.768	7.228	666.7E6	223.2E6	70.027	67.846
8) T	DICHLORPROP	7.080	7.538	1773.2E6	543.3E6	691.216	690.603
9) T	2,4-D	7.277	7.821	2137.2E6	679.7E6	703.636	697.770
10) T	Pentachlo...	7.506	8.243	25829.1E6	8092.4E6	736.637	713.901
11) T	2,4,5-TP ...	8.027	8.601	10800.8E6	3092.9E6	710.993	701.329
12) T	2,4,5-T	8.282	8.970	10372.4E6	3121.5E6	718.094	723.487
13) T	2,4-DB	8.786	9.473	1685.5E6	398.8E6	725.489	720.570
14) T	DINOSEB	9.838	9.800	7706.2E6	2177.4E6	709.734	710.207
15) T	Picloram	9.665	10.749	13818.1E6	4554.3E6	709.137	695.811
16) T	DCPA	10.104	10.720	13241.3E6	3934.0E6	705.711	725.838

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122523\
Data File : PS025208.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2023 02:21
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 12/26/2023
Supervised By : Ankita Jodhani 12/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 03:36:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS112823.M
Quant Title : 8080.M
QLast Update : Tue Nov 28 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

