

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS122619.M
 Title : 8080.M
 Last Update : Fri Dec 27 02:43:58 2019
 Response Via : Initial Calibration

Calibration Files

750 =PS008154.D 1000 =PS008155.D 1500 =PS008156.D
 500 =PS008153.D 200 =PS008152.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	2.378	2.382	2.425	2.372	2.446	2.401	E6 1.36
2) T	3,5-DICHLOROBENZOI...	2.536	2.395	2.496	2.416	2.578	2.484	E6 3.13
3) T	4-Nitrophenol	7.220	7.932	7.791	6.165	5.061	6.834	E5 17.71
4) S	2,4-DCAA	2.012	1.882	1.962	1.949	2.297	2.020	E6 7.99
5) T	DICAMBA	7.349	6.820	6.971	7.104	8.114	7.272	E6 7.01
6) T	MCPP	5.590	5.451	5.969	5.105	5.155	5.454	E6 6.45
7) T	MCPA	7.803	7.563	8.154	7.230	7.489	7.648	E6 4.56
8) T	DICHLORPROP	1.789	1.671	1.762	1.703	1.935	1.772	E6 5.78
9) T	2,4-D	1.678	1.614	1.774	1.497	1.642	1.641	E6 6.14
10) T	Pentachlorophenol	2.643	2.388	2.311	2.654	3.099	2.619	E7 11.77
11) T	2,4,5-TP (SILVEX)	1.044	0.962	0.976	1.017	1.158	1.031	E7 7.54
12) T	2,4,5-T	1.176	1.076	1.113	1.261	1.383	1.202	E7 10.25
13) T	2,4-DB	1.343	1.309	1.506	1.298	1.188	1.329	E6 8.64
14) T	DINOSEB	0.949	0.869	0.851	0.992	1.184	0.969	E7 13.78
15) T	Picloram	1.162	1.243	1.365	0.995	0.665	1.086	E7 24.94
16) T	DCPA	1.333	1.299	1.249	1.375	1.528	1.357	E7 7.81

Signal #2 Calibration Files

750 =PS008154.D 1000 =PS008155.D 1500 =PS008156.D
 500 =PS008153.D 200 =PS008152.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	1.012	1.034	1.090	0.975	0.989	1.020	E6 4.43
2) T	3,5-DICHLOROBENZOI...	1.173	1.168	1.159	1.145	1.176	1.164	E6 1.08
3) T	4-Nitrophenol	4.853	5.270	5.221	4.847	4.577	4.954	E5 5.84
4) S	2,4-DCAA	8.853	8.850	8.987	8.878	9.202	8.954	E5 1.67
5) T	DICAMBA	3.390	3.340	3.275	3.441	3.588	3.407	E6 3.47
6) T	MCPP	2.585	2.641	2.724	2.539	2.451	2.588	E6 3.99
7) T	MCPA	3.727	3.755	3.857	3.725	3.958	3.804	E6 2.66
8) T	DICHLORPROP	8.485	8.444	8.531	8.557	8.785	8.560	E5 1.55
9) T	2,4-D	1.001	0.999	1.015	0.999	1.063	1.015	E6 2.69
10) T	Pentachlorophenol	1.258	1.192	1.099	1.335	1.479	1.273	E7 11.35
11) T	2,4,5-TP (SILVEX)	4.967	4.855	4.745	5.059	5.359	4.997	E6 4.68
12) T	2,4,5-T	4.249	4.255	4.308	4.154	3.932	4.180	E6 3.57
13) T	2,4-DB	6.195	6.288	6.743	6.060	5.993	6.256	E5 4.73
14) T	DINOSEB	3.720	3.651	3.651	3.744	3.961	3.745	E6 3.40
15) T	Picloram	6.789	6.934	7.070	6.341	5.498	6.527	E6 9.76
16) T	DCPA	6.559	6.498	6.198	6.785	7.316	6.671	E6 6.25

(#) = Out of Range